

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102318\
 Data File : PL041275.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Oct 2018 22:02
 Operator : AJ\SJ
 Sample : MDL-TOX-W-02
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MDL-TOX-W-02

Manual Integrations
 APPROVED

Sohil
 10/25/2018 8:37:12 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 08:01:34 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102318CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 24 01:06:36 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.475	4.103	103.9E6	55508062	16.261	17.178
27)	SA Decachlor...	8.222	9.233	228.6E6	100.1E6	35.627	37.329
Target Compounds							
22)	Toxaphene-1	5.727	6.513	3291853	1466581	64.206m	100.798m#
23)	Toxaphene-2	5.888	6.806	3528295	1448027	68.738	86.795m#
24)	Toxaphene-3	6.330	7.215	7512165	2381616	63.599m	86.597m#
25)	Toxaphene-4	6.575	7.473	8957293	1622598	68.775m	87.187 #
26)	Toxaphene-5	7.063	7.795	9489458	2839089	70.819	89.229 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102318\
Data File : PL041275.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Oct 2018 22:02
Operator : AJ\SJ
Sample : MDL-TOX-W-02
Misc :
ALS Vial : 31 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 08:01:34 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102318CLP.M
Quant Title : GC Extractables
QLast Update : Wed Oct 24 01:06:36 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



232
Sohil
10/25/2018 8:37:12 AM

Manual Integrations
APPROVED

Instrument :
ECD_L
Client Sample :
MDL-TOX-W-02