

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102423\
 Data File : PL086125.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2023 18:01
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 18:12:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 04 05:29:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.355	2.621	142.3E6	44717200	47.848	47.871
28)	SA Decachlor...	8.854	7.772	125.3E6	52878781	57.520	57.291
Target Compounds							
2)	A alpha-BHC	3.809	3.120	209.8E6	61680944	48.882	49.026
3)	MA gamma-BHC...	4.141	3.448	203.9E6	61012206	49.323	49.469
4)	MA Heptachlor	4.727	3.783	182.8E6	56741188	47.252	48.786
5)	MB Aldrin	5.067	4.060	192.0E6	58334450	49.536	52.165
6)	B beta-BHC	4.344	3.750	84312018	27032495	50.986	51.927
7)	B delta-BHC	4.587	3.975	185.0E6	60267549	48.243	52.454
8)	B Heptachlo...	5.497	4.565	167.1E6	54341026	48.443	54.103
9)	A Endosulfan I	5.881	4.932	165.0E6	51219507	51.022	47.746
10)	B gamma-Chl...	5.754	4.816	176.3E6	56722447	51.533	57.273
11)	B alpha-Chl...	5.833	4.880	174.4E6	56457491	51.021	51.761
12)	B 4,4'-DDE	6.017	5.079	152.2E6	48919472	54.413	59.575
13)	MA Dieldrin	6.157	5.198	164.3E6	54344628	49.996	56.013
14)	MA Endrin	6.386	5.472	145.9E6	48811128	51.192	54.295
15)	B Endosulfa...	6.609	5.769	144.0E6	49805340	52.272	52.375
16)	A 4,4'-DDD	6.536	5.635	116.6E6	40637184	57.896	58.383
17)	MA 4,4'-DDT	6.851	5.886	111.9E6	40382816	47.467	51.069
18)	B Endrin al...	6.739	5.951	102.6E6	39021925	50.619	52.329
19)	B Endosulfa...	6.976	6.175	133.5E6	48945275	53.199	54.256
20)	A Methoxychlor	7.336	6.470	62924940	24831302	45.055	47.201
21)	B Endrin ke...	7.458	6.676	138.9E6	57925444	48.238	50.396
22)	Mirex	7.930	6.853	112.9E6	47998116	50.997	50.890

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102423\
Data File : PL086125.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2023 18:01
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 18:12:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
Quant Title : GC Extractables
QLast Update : Wed Oct 04 05:29:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

