

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102423\
 Data File : PL086136.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2023 20:33
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 21:25:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 04 05:29:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.355	2.621	145.8E6	44948615	49.017	48.119
28)	SA Decachlor...	8.854	7.772	129.8E6	53269899	59.595	57.715
Target Compounds							
2)	A alpha-BHC	3.809	3.120	215.4E6	61774725	50.195	49.100
3)	MA gamma-BHC...	4.140	3.447	208.3E6	61775903	50.404	50.088
4)	MA Heptachlor	4.727	3.783	189.3E6	57266509	48.953	49.238
5)	MB Aldrin	5.067	4.060	195.7E6	58473540	50.491	52.290
6)	B beta-BHC	4.343	3.749	86452023	27151993	52.280	52.156
7)	B delta-BHC	4.586	3.974	196.2E6	60939781	51.174	53.039
8)	B Heptachlo...	5.496	4.564	170.4E6	54033908	49.406	53.797
9)	A Endosulfan I	5.881	4.931	163.8E6	51011842	50.636	47.553
10)	B gamma-Chl...	5.754	4.816	175.8E6	56357077	51.368	56.904
11)	B alpha-Chl...	5.834	4.879	174.7E6	55962832	51.125	51.308
12)	B 4,4'-DDE	6.017	5.079	153.0E6	48094482	54.703	58.571
13)	MA Dieldrin	6.158	5.197	167.5E6	53967404	50.987	55.624
14)	MA Endrin	6.386	5.471	147.6E6	48710034	51.805	54.182
15)	B Endosulfa...	6.608	5.768	145.9E6	49190901	52.959	51.729
16)	A 4,4'-DDD	6.535	5.635	118.5E6	40382592	58.828	58.017
17)	MA 4,4'-DDT	6.850	5.885	114.6E6	40376241	48.607	51.060
18)	B Endrin al...	6.739	5.950	105.4E6	38848176	52.011	52.096
19)	B Endosulfa...	6.975	6.174	136.3E6	49004983	54.289	54.322
20)	A Methoxychlor	7.336	6.470	66936534	25343899	47.928	48.175
21)	B Endrin ke...	7.457	6.676	148.4E6	57888425	51.538	50.363
22)	Mirex	7.930	6.853	113.9E6	48368082	51.455	51.282

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102423\
Data File : PL086136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2023 20:33
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 21:25:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100323.M
Quant Title : GC Extractables
QLast Update : Wed Oct 04 05:29:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

