

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102519\
 Data File : PL053754.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2019 16:37
 Operator : SG\AJ
 Sample : K5514-01MS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ROLL-OFF-WATERMS

Manual Integrations
 APPROVED

Ankita
 10/28/2019 12:29:02 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 22:33:24 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	4.045	186.5E6	44160863	17.209	17.648
28)	SA Decachlor...	8.327	9.121	146.1E6	44736843	14.886	15.887
Target Compounds							
2)	A alpha-BHC	3.986	4.436	929.2E6	194.5E6	53.288	56.268
3)	MA gamma-BHC...	4.272	4.721	852.1E6	181.4E6	54.715	56.317
4)	MA Heptachlor	4.579	5.230	881.8E6	176.7E6	55.868	54.852
5)	MB Aldrin	4.830	5.536	802.9E6	163.8E6	56.162	54.922
6)	B beta-BHC	4.522	4.891	353.7E6	75403042	53.667	48.816
7)	B delta-BHC	4.731	5.108	766.9E6	138.1E6	50.383	45.721
8)	B Heptachlor...	5.276	5.920	765.9E6	147.2E6	56.738	50.473m
9)	A Endosulfan I	5.618	6.279	691.6E6	148.4E6	56.260	54.731
10)	B gamma-Chl...	5.504	6.154	765.7E6	158.4E6	56.120	54.293m
11)	B alpha-Chl...	5.563	6.228	754.2E6	158.6E6	56.692	54.332
12)	B 4,4'-DDE	5.727	6.385	675.9E6	154.5E6	56.885	54.842
13)	MA Dieldrin	5.863	6.535	731.8E6	159.4E6	56.149	55.844
14)	MA Endrin	6.121	6.753	680.9E6	138.6E6	58.746	57.114
15)	B Endosulfa...	6.395	6.962	613.4E6	139.1E6	54.860	54.444
16)	A 4,4'-DDD	6.244	6.874	514.3E6	120.6E6	57.785	57.862
17)	MA 4,4'-DDT	6.487	7.173	567.5E6	132.3E6	57.283	57.216
18)	B Endrin al...	6.563	7.086	483.3E6	118.2E6	52.707	53.793
19)	B Endosulfa...	6.776	7.309	559.8E6	129.7E6	55.031	52.854
20)	A Methoxychlor	7.032	7.630	282.1E6	69125992	57.211	53.218
21)	B Endrin ke...	7.270	7.780	599.1E6	140.5E6	55.655	54.501
22)	Mirex	7.463	8.235	535.4E6	120.5E6	62.597	59.544

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102519\
Data File : PL053754.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2019 16:37
Operator : SG\AJ
Sample : K5514-01MS
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
ROLL-OFF-WATERMS

Manual Integrations
APPROVED

Ankita
10/28/2019 12:29:02 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 22:33:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

