

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102522\
 Data File : PL078547.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2022 20:28
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 21:57:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.322	2.571	94584106	127.2E6	49.598	49.829
28)	SA Decachlor...	8.772	7.640	76279481	92249098	50.706	48.703
Target Compounds							
2)	A alpha-BHC	3.782	3.058	147.5E6	198.0E6	51.380	50.148
3)	MA gamma-BHC...	4.115	3.379	137.1E6	181.0E6	51.317	50.351
4)	MA Heptachlor	4.686	3.706	131.2E6	168.7E6	50.533	48.646
5)	MB Aldrin	5.024	3.978	128.3E6	176.0E6	51.785	50.267
6)	B beta-BHC	4.329	3.679	58549949	70421435	50.826	49.080
7)	B delta-BHC	4.570	3.899	126.5E6	179.1E6	55.577	49.192
8)	B Heptachlo...	5.457	4.475	116.8E6	149.1E6	55.026	49.473
9)	A Endosulfan I	5.842	4.837	107.8E6	140.1E6	50.540	48.953
10)	B gamma-Chl...	5.713	4.722	117.6E6	149.0E6	52.083	48.622
11)	B alpha-Chl...	5.789	4.784	115.9E6	145.4E6	50.087	48.774
12)	B 4,4'-DDE	5.972	4.979	102.3E6	136.2E6	51.114	50.016
13)	MA Dieldrin	6.119	5.099	108.6E6	145.3E6	50.367	49.053
14)	MA Endrin	6.349	5.370	86527118	120.5E6	45.593	46.262
15)	B Endosulfa...	6.579	5.666	87352567	116.7E6	46.559	51.249
16)	A 4,4'-DDD	6.497	5.530	86551586	116.5E6	51.920	53.035
17)	MA 4,4'-DDT	6.806	5.777	79926851	103.3E6	44.594	47.606
18)	B Endrin al...	6.711	5.845	71645814	91716533	49.239	48.869
19)	B Endosulfa...	6.943	6.067	89570958	117.3E6	48.682	49.278
20)	A Methoxychlor	7.291	6.354	41326710	52359009	44.877	46.140
21)	B Endrin ke...	7.429	6.563	97206537	135.6E6	51.667	52.470
22)	Mirex	7.887	6.736	73160710	94874293	50.116	48.209

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102522\
Data File : PL078547.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2022 20:28
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 21:57:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

