

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102620\
 Data File : PL062579.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Oct 2020 10:30
 Operator : DD\AJ
 Sample : L4486-02DL 50X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MS-SOILDL

Manual Integrations
 APPROVED

Ankita
 10/27/2020 10:05:15 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 04:01:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
16) A	4,4'-DDD	6.064	6.768	7875864	14397196	5.568 5.624
17) MA	4,4'-DDT	6.302	7.067	100.1E6	162.6E6	74.974m 72.427

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102620\
Data File : PL062579.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Oct 2020 10:30
Operator : DD\AJ
Sample : L4486-02DL 50X
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
MS-SOILDL


Manual Integrations
APPROVED

Ankita
10/27/2020 10:05:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 04:01:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 3.412 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
MS-SOILD


Manual Integrations
APPROVED

Ankita
10/27/2020 10:05:15 AM




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 3.952 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 6.064 min
Delta R.T.: 0.002 min
Response: 7875864
Conc: 5.57 ng/ml



#16 4,4'-DDD

R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 14397196
Conc: 5.62 ng/ml




#17 4,4'-DDT

R.T.: 6.302 min
Delta R.T.: 0.001 min
Response: 100057458
Conc: 74.97 ng/ml m

Instrument :
ECD_L
ClientSampleId :
MS-SOILD


Manual Integrations
APPROVED

Ankita
10/27/2020 10:05:15 AM




#17 4,4'-DDT

R.T.: 7.067 min
Delta R.T.: 0.000 min
Response: 162618292
Conc: 72.43 ng/ml



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.126 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.989 min
Response: 0
Conc: N.D.