

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102620\
 Data File : PL062585.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Oct 2020 12:11
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 04:03:36 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.413	3.952	89302707	159.6E6	51.918	54.317
28)	SA Decachlor...	8.127	8.990	76945032	105.6E6	47.962	50.475
Target Compounds							
2)	A alpha-BHC	3.844	4.338	132.9E6	244.7E6	52.136	55.073
3)	MA gamma-BHC...	4.124	4.621	118.5E6	226.4E6	51.164	54.228
4)	MA Heptachlor	4.421	5.126	108.2E6	209.9E6	50.445	52.642
5)	MB Aldrin	4.665	5.429	98245398	195.8E6	51.471	54.106
6)	B beta-BHC	4.373	4.792	49311055	97389655	51.624	51.720
7)	B delta-BHC	4.574	5.007	111.9E6	200.8E6	51.488	57.372
8)	B Heptachlo...	5.104	5.814	93330836	176.3E6	51.152	54.429
9)	A Endosulfan I	5.439	6.170	85109488	158.2E6	50.915	50.159
10)	B gamma-Chl...	5.329	6.048	91421295	174.8E6	50.745	51.913
11)	B alpha-Chl...	5.387	6.120	90943247	172.8E6	50.307	51.304
12)	B 4,4'-DDE	5.552	6.279	84615035	163.9E6	51.144	50.857
13)	MA Dieldrin	5.680	6.426	89805582	161.0E6	50.825	50.284
14)	MA Endrin	5.935	6.643	78866638	130.0E6	47.402	46.586
15)	B Endosulfa...	6.207	6.853	81767930	135.0E6	49.597	47.238
16)	A 4,4'-DDD	6.064	6.768	75912776	133.4E6	53.668	52.098
17)	MA 4,4'-DDT	6.303	7.067	60555014	104.5E6	45.374	46.532
18)	B Endrin al...	6.374	6.977	70186724	114.3E6	48.335	50.122
19)	B Endosulfa...	6.586	7.199	80234803	124.8E6	48.620	52.004
20)	A Methoxychlor	6.847	7.526	37791108	58028572	43.877	46.967
21)	B Endrin ke...	7.072	7.668	93334971	130.3E6	51.766	54.971
22)	Mirex	7.258	8.121	68390119	88225986	48.241	51.773


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102620\
Data File : PL062585.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Oct 2020 12:11
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 04:03:36 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.413 min
Delta R.T.: 0.000 min
Response: 89302707
Conc: 51.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 3.952 min
Delta R.T.: 0.000 min
Response: 159621416
Conc: 54.32 ng/ml




#2 alpha-BHC

R.T.: 3.844 min
Delta R.T.: 0.000 min
Response: 132931623
Conc: 52.14 ng/ml



#2 alpha-BHC


R.T.: 4.338 min
Delta R.T.: 0.000 min
Response: 244701825
Conc: 55.07 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.124 min
Delta R.T.: 0.000 min
Response: 118531465
Conc: 51.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 4.621 min
Delta R.T.: 0.000 min
Response: 226442951
Conc: 54.23 ng/ml




#4 Heptachlor

R.T.: 4.421 min
Delta R.T.: 0.000 min
Response: 108244797
Conc: 50.45 ng/ml



#4 Heptachlor

R.T.: 5.126 min
Delta R.T.: 0.000 min
Response: 209930031
Conc: 52.64 ng/ml



#5 Aldrin


R.T.: 4.665 min

Delta R.T.: 0.000 min

Response: 98245398

Conc: 51.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 5.429 min

Delta R.T.: 0.000 min

Response: 195835395

Conc: 54.11 ng/ml




#6 beta-BHC

R.T.: 4.373 min

Delta R.T.: 0.000 min

Response: 49311055

Conc: 51.62 ng/ml




#6 beta-BHC

R.T.: 4.792 min

Delta R.T.: 0.000 min

Response: 97389655


Conc: 51.72 ng/ml



#7 delta-BHC


R.T.: 4.574 min
Delta R.T.: 0.000 min
Response: 111894101
Conc: 51.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 5.007 min
Delta R.T.: 0.000 min
Response: 200770075
Conc: 57.37 ng/ml




#8 Heptachlor epoxide

R.T.: 5.104 min
Delta R.T.: 0.000 min
Response: 93330836
Conc: 51.15 ng/ml



#8 Heptachlor epoxide


R.T.: 5.814 min
Delta R.T.: 0.000 min
Response: 176275209
Conc: 54.43 ng/ml



#9 Endosulfan I


R.T.: 5.439 min
Delta R.T.: 0.000 min
Response: 85109488
Conc: 50.91 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 6.170 min
Delta R.T.: 0.000 min
Response: 158193675
Conc: 50.16 ng/ml




#10 gamma-Chlordane

R.T.: 5.329 min
Delta R.T.: 0.000 min
Response: 91421295
Conc: 50.75 ng/ml



#10 gamma-Chlordane


R.T.: 6.048 min
Delta R.T.: 0.000 min
Response: 174756030
Conc: 51.91 ng/ml



#11 alpha-Chlordane


R.T.: 5.387 min
Delta R.T.: 0.000 min
Response: 90943247
Conc: 50.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 6.120 min
Delta R.T.: 0.000 min
Response: 172838110
Conc: 51.30 ng/ml




#12 4,4'-DDE

R.T.: 5.552 min
Delta R.T.: 0.001 min
Response: 84615035
Conc: 51.14 ng/ml



#12 4,4'-DDE


R.T.: 6.279 min
Delta R.T.: 0.000 min
Response: 163861480
Conc: 50.86 ng/ml



#13 Dieldrin


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 89805582
Conc: 50.82 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 6.426 min
Delta R.T.: 0.000 min
Response: 160952306
Conc: 50.28 ng/ml




#14 Endrin

R.T.: 5.935 min
Delta R.T.: 0.002 min
Response: 78866638
Conc: 47.40 ng/ml



#14 Endrin


R.T.: 6.643 min
Delta R.T.: 0.000 min
Response: 130025820
Conc: 46.59 ng/ml



#15 Endosulfan II


R.T.: 6.207 min
Delta R.T.: 0.002 min
Response: 81767930
Conc: 49.60 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.853 min
Delta R.T.: 0.000 min
Response: 135038021
Conc: 47.24 ng/ml




#16 4,4'-DDD

R.T.: 6.064 min
Delta R.T.: 0.002 min
Response: 75912776
Conc: 53.67 ng/ml



#16 4,4'-DDD


R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 133368853
Conc: 52.10 ng/ml



#17 4,4'-DDT


R.T.: 6.303 min
Delta R.T.: 0.002 min
Response: 60555014
Conc: 45.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 7.067 min
Delta R.T.: 0.000 min
Response: 104477048
Conc: 46.53 ng/ml




#18 Endrin aldehyde

R.T.: 6.374 min
Delta R.T.: 0.002 min
Response: 70186724
Conc: 48.34 ng/ml



#18 Endrin aldehyde


R.T.: 6.977 min
Delta R.T.: 0.000 min
Response: 114308222
Conc: 50.12 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.586 min
Delta R.T.: 0.001 min
Response: 80234803
Conc: 48.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 7.199 min
Delta R.T.: 0.000 min
Response: 124769257
Conc: 52.00 ng/ml




#20 Methoxychlor

R.T.: 6.847 min
Delta R.T.: 0.002 min
Response: 37791108
Conc: 43.88 ng/ml



#20 Methoxychlor


R.T.: 7.526 min
Delta R.T.: 0.000 min
Response: 58028572
Conc: 46.97 ng/ml



#21 Endrin ketone


R.T.: 7.072 min
Delta R.T.: 0.000 min
Response: 93334971
Conc: 51.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 7.668 min
Delta R.T.: 0.000 min
Response: 130269661
Conc: 54.97 ng/ml




#22 Mirex

R.T.: 7.258 min
Delta R.T.: 0.001 min
Response: 68390119
Conc: 48.24 ng/ml



#22 Mirex


R.T.: 8.121 min
Delta R.T.: 0.000 min
Response: 88225986
Conc: 51.77 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.127 min
Delta R.T.: 0.001 min
Response: 76945032
Conc: 47.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 8.990 min
Delta R.T.: 0.000 min
Response: 105641118
Conc: 50.47 ng/ml