

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102621\
 Data File : PL070565.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2021 07:48
 Operator : AR\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 08:13:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102621.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 27 07:49:52 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	50.000	56.156	-12.3	111	0.00
23	Chlordane-1	500.000	538.256	-7.7	106	0.00
24	Chlordane-2	500.000	534.282	-6.9	106	0.00
25	Chlordane-3	500.000	539.980	-8.0	106	0.00
26	Chlordane-4	500.000	544.234	-8.8	107	0.00
27	Chlordane-5	500.000	533.022	-6.6	105	0.00
28 SA	Decachlorobiphenyl	50.000	52.425	-4.8	108	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	50.000	54.495	-9.0	107	0.00
23	Chlordane-1	500.000	549.173	-9.8	105	0.00
24	Chlordane-2	500.000	526.271	-5.3	103	0.00
25	Chlordane-3	500.000	537.895	-7.6	104	0.00
26	Chlordane-4	500.000	540.265	-8.1	104	0.00
27	Chlordane-5	500.000	540.856	-8.2	105	0.00
28 SA	Decachlorobiphenyl	50.000	54.925	-9.8	109	0.00

Evaluate Continuing Calibration Report - Not Found

2 A	alpha-BHC	50.000	0.000	100.0#	0	-5.33#
3 MA	gamma-BHC (Lindane)	50.000	0.000	100.0#	0	-5.75#
4 MA	Heptachlor	50.000	0.000	100.0#	0	-6.16#
5 MB	Aldrin	50.000	0.000	100.0#	0	-6.49#
6 B	beta-BHC	50.000	0.000	100.0#	0	-6.13#
7 B	delta-BHC	50.000	0.000	100.0#	0	-6.39#
8 B	Heptachlor epoxide	50.000	0.000	100.0#	0	-7.06#
9 A	Endosulfan I	50.000	0.000	100.0#	0	-7.47#
10 B	gamma-Chlordane	50.000	0.000	100.0#	0	-7.34#
11 B	alpha-Chlordane	50.000	0.000	100.0#	0	-7.41#
12 B	4,4'-DDE	50.000	0.000	100.0#	0	-7.62#
13 MA	Dieldrin	50.000	0.000	100.0#	0	-7.75#
14 MA	Endrin	50.000	0.000	100.0#	0	-8.04#
15 B	Endosulfan II	50.000	0.000	100.0#	0	-8.35#
16 A	4,4'-DDD	50.000	0.000	100.0#	0	-8.21#
17 MA	4,4'-DDT	50.000	0.000	100.0#	0	-8.47#
18 B	Endrin aldehyde	50.000	0.000	100.0#	0	-8.54#
19 B	Endosulfan Sulfate	50.000	0.000	100.0#	0	-8.77#
20 A	Methoxychlor	50.000	0.000	100.0#	0	-9.06#
21 B	Endrin ketone	50.000	0.000	100.0#	0	-9.29#
22	Mirex	50.000	0.000	100.0#	0	-9.48#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102621\
 Data File : PL070565.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2021 07:48
 Operator : AR\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 08:13:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102621.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 27 07:49:52 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)

Signal #2					
2 A alpha-BHC	50.000	0.000	100.0#	0	-6.14#
3 MA gamma-BHC (Lindane)	50.000	0.000	100.0#	0	-6.54#
4 MA Heptachlor	50.000	0.000	100.0#	0	-7.19#
5 MB Aldrin	50.000	0.000	100.0#	0	-7.56#
6 B beta-BHC	50.000	0.000	100.0#	0	-6.79#
7 B delta-BHC	50.000	0.000	100.0#	0	-7.06#
8 B Heptachlor epoxide	50.000	0.000	100.0#	0	-8.03#
9 A Endosulfan I	50.000	0.000	100.0#	0	-8.43#
10 B gamma-Chlordane	50.000	0.000	100.0#	0	-8.30#
11 B alpha-Chlordane	50.000	0.000	100.0#	0	-8.38#
12 B 4,4'-DDE	50.000	0.000	100.0#	0	-8.57#
13 MA Dieldrin	50.000	0.000	100.0#	0	-8.72#
14 MA Endrin	50.000	0.000	100.0#	0	-8.96#
15 B Endosulfan II	50.000	0.000	100.0#	0	-9.20#
16 A 4,4'-DDD	50.000	0.000	100.0#	0	-9.11#
17 MA 4,4'-DDT	50.000	0.000	100.0#	0	-9.42#
18 B Endrin aldehyde	50.000	0.000	100.0#	0	-9.33#
19 B Endosulfan Sulfate	50.000	0.000	100.0#	0	-9.57#
20 A Methoxychlor	50.000	0.000	100.0#	0	-9.91#
21 B Endrin ketone	50.000	0.000	100.0#	0	-10.07#
22 Mirex	50.000	0.000	100.0#	0	-10.53#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0