

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102720\
 Data File : PL062625.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2020 16:04
 Operator : DD\AJ
 Sample : L4516-04
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 05:05:05 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.412	3.951	34194351	58702459	19.879	19.976
28)	SA Decachlor...	8.127	8.990	31814526	44427027	19.831	21.227
Target Compounds							
9)	A Endosulfan I	0.000	6.171	0	6749196	N.D.	2.140 #
14)	MA Endrin	0.000	6.666f	0	6254935	N.D.	2.241 #
15)	B Endosulfa...	0.000	6.868f	0	2217369	N.D.	0.776 #
17)	MA 4,4'-DDT	6.317f	0.000	1262951	0	0.946	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102720\
Data File : PL062625.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2020 16:04
Operator : DD\AJ
Sample : L4516-04
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 05:05:05 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.412 min
Delta R.T.: 0.000 min
Response: 34194351
Conc: 19.88 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.951 min
Delta R.T.: 0.000 min
Response: 58702459
Conc: 19.98 ng/ml




#9 Endosulfan I

R.T.: 0.000 min
Exp R.T. : 5.438 min
Response: 0
Conc: N.D.



#9 Endosulfan I


R.T.: 6.171 min
Delta R.T.: 0.000 min
Response: 6749196
Conc: 2.14 ng/ml



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 5.933 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 6.666 min
Delta R.T.: 0.022 min
Response: 6254935
Conc: 2.24 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 6.205 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.868 min
Delta R.T.: 0.015 min
Response: 2217369
Conc: 0.78 ng/ml



#17 4,4'-DDT


R.T.: 6.317 min
Delta R.T.: 0.016 min
Response: 1262951
Conc: 0.95 ng/ml

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T. : 7.066 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.127 min
Delta R.T.: 0.000 min
Response: 31814526
Conc: 19.83 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.990 min
Delta R.T.: 0.000 min
Response: 44427027
Conc: 21.23 ng/ml