

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102720\
 Data File : PL062620.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2020 14:47
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 05:04:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.413	3.951	84428643	152.0E6	49.084	51.724
28)	SA Decachlor...	8.126	8.990	81870881	114.6E6	51.033	54.767
Target Compounds							
2)	A alpha-BHC	3.844	4.337	126.0E6	234.4E6	49.403	52.748
3)	MA gamma-BHC...	4.124	4.620	111.9E6	217.0E6	48.315	51.979
4)	MA Heptachlor	4.421	5.126	103.4E6	202.5E6	48.207	50.781
5)	MB Aldrin	4.665	5.428	93294016	189.2E6	48.877	52.268
6)	B beta-BHC	4.373	4.792	46919784	93475992	49.121	49.642
7)	B delta-BHC	4.574	5.006	106.4E6	198.8E6	48.943	56.817
8)	B Heptachlo...	5.104	5.814	90293334	174.7E6	49.487	53.931
9)	A Endosulfan I	5.439	6.169	82909569	157.8E6	49.599	50.028
10)	B gamma-Chl...	5.329	6.047	88137922	172.1E6	48.923	51.120
11)	B alpha-Chl...	5.387	6.120	87968021	170.8E6	48.661	50.685
12)	B 4,4'-DDE	5.551	6.279	80831333	162.9E6	48.857	50.545
13)	MA Dieldrin	5.680	6.425	86487703	161.5E6	48.947	50.468
14)	MA Endrin	5.934	6.643	75131300	129.2E6	45.157	46.287
15)	B Endosulfa...	6.207	6.852	80240610	137.1E6	48.671	47.953
16)	A 4,4'-DDD	6.064	6.768	74303441	134.8E6	52.530	52.670
17)	MA 4,4'-DDT	6.302	7.067	58669406	106.5E6	43.961	47.436
18)	B Endrin al...	6.374	6.977	69883927	117.8E6	48.127	51.652
19)	B Endosulfa...	6.585	7.199	79926215	128.1E6	48.433	53.380
20)	A Methoxychlor	6.846	7.526	37223854	59559888	43.219	48.207
21)	B Endrin ke...	7.072	7.668	94560142	136.7E6	52.446	57.691
22)	Mirex	7.258	8.122	68784917	93409723	48.519	54.815

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102720\
Data File : PL062620.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2020 14:47
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 05:04:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

