

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102722\
 Data File : PL078566.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Oct 2022 18:27
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/27/2022
 Supervised By :Ankita Jodhani 10/27/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 26 21:44:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.322	2.571	102.1E6	137.4E6	53.517	53.818
28)	SA Decachlor...	8.776	7.642	79219046	99305076	52.660	52.429
Target Compounds							
2)	A alpha-BHC	3.783	3.059	161.2E6	216.2E6	56.150	54.737
3)	MA gamma-BHC...	4.117	3.380	149.9E6	198.5E6	56.088	55.238
4)	MA Heptachlor	4.687	3.707	149.4E6	193.2E6	57.556	55.725
5)	MB Aldrin	5.026	3.978	139.8E6	187.4E6	56.462	53.523m
6)	B beta-BHC	4.331	3.680	63204157	76803818	54.866	53.528
7)	B delta-BHC	4.572	3.900	131.6E6	199.5E6	57.801	54.790
8)	B Heptachlo...	5.458	4.476	128.5E6	163.8E6	60.527m	54.367
9)	A Endosulfan I	5.844	4.839	117.0E6	152.8E6	54.851	53.388
10)	B gamma-Chl...	5.715	4.724	125.6E6	165.2E6	55.641	53.913
11)	B alpha-Chl...	5.791	4.786	124.8E6	161.8E6	53.959	54.281
12)	B 4,4'-DDE	5.974	4.980	110.6E6	150.7E6	55.244	55.355
13)	MA Dieldrin	6.122	5.101	118.9E6	165.0E6	55.138	55.681
14)	MA Endrin	6.351	5.372	110.5E6	145.1E6	58.233	55.707
15)	B Endosulfa...	6.580	5.668	93749009	126.9E6	49.968	55.734
16)	A 4,4'-DDD	6.499	5.532	88396489	123.6E6	53.026	56.262
17)	MA 4,4'-DDT	6.809	5.778	98469839	128.7E6	54.940	59.325
18)	B Endrin al...	6.713	5.846	76425965	99589727	52.524	53.064
19)	B Endosulfa...	6.946	6.069	96473003	129.6E6	52.433	54.442
20)	A Methoxychlor	7.294	6.355	50554327	67211368	54.897	59.228
21)	B Endrin ke...	7.432	6.565	99616280	147.3E6	52.948	56.983
22)	Mirex	7.890	6.738	76741966	104.9E6	52.570	53.293

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102722\
Data File : PL078566.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Oct 2022 18:27
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/27/2022
Supervised By : Ankita Jodhani 10/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 21:44:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

