

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102723\
 Data File : PL086189.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2023 11:05
 Operator : AR\AJ
 Sample : PB156294BL
 Misc : WATER
 ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 05:23:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102623CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 27 01:35:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.356	2.621	59593908	18441309	20.909	20.332
27)	SA Decachlor...	8.858	7.773	55890851	26019736	22.427	22.667
Target Compounds							
2)	A alpha-BHC	3.845f	0.000	82646	0	<MDL	N.D. #
3)	MA gamma-BHC...	4.131	0.000	225480	0	<MDL	N.D. #
5)	MB Aldrin	5.057	0.000	70024	0	<MDL	N.D. #
6)	B beta-BHC	4.363	0.000	2256255	0	1.218	N.D. #
7)	B delta-BHC	4.594	0.000	1604179	0	0.431	N.D. #
8)	B Heptachlo...	5.428f	0.000	15115857	0	4.458	N.D. #
12)	B 4,4'-DDE	6.012	0.000	282866	0	<MDL	N.D. #
14)	MA Endrin	6.388	0.000	75196	0	<MDL	N.D. #
15)	B Endosulfa...	6.610	0.000	3970356	0	1.269	N.D. #
18)	B Endrin al...	6.671f	0.000	633619	0	0.272	N.D. #
19)	B Endosulfa...	6.962	0.000	372867	0	0.131	N.D. #
20)	A Methoxychlor	7.398f	0.000	6178826	0	4.899	N.D. #
21)	B Endrin ke...	7.514f	0.000	1163494	0	0.404	N.D. #
22)	Toxaphene-1	6.012f	0.000	282866	0	12.544	N.D. #
24)	Toxaphene-3	6.962	0.000	372867	0	5.941	N.D. #
25)	Toxaphene-4	7.132f	0.000	95678	0	2.546	N.D. #
26)	Toxaphene-5	7.722	0.000	221044	0	5.009	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102723\
Data File : PL086189.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2023 11:05
Operator : AR\AJ
Sample : PB156294BL
Misc : WATER
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 05:23:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102623CLP.M
Quant Title : GC Extractables
QLast Update : Fri Oct 27 01:35:51 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.356 min
Delta R.T.: 0.002 min
Response: 59593908
Conc: 20.91 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.621 min
Delta R.T.: 0.000 min
Response: 18441309
Conc: 20.33 ng/ml




#6 beta-BHC

R.T.: 4.363 min
Delta R.T.: 0.019 min
Response: 2256255
Conc: 1.22 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T.: 3.749 min
Response: 0
Conc: N.D.




#7 delta-BHC

R.T.: 4.594 min
Delta R.T.: 0.007 min
Response: 1604179
Conc: 0.43 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T.: 3.974 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.428 min
Delta R.T.: -0.069 min
Response: 15115857
Conc: 4.46 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T.: 4.564 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 6.610 min
Delta R.T.: 0.002 min
Response: 3970356
Conc: 1.27 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 5.769 min
Response: 0
Conc: N.D.




#18 Endrin aldehyde

R.T.: 6.671 min
Delta R.T.: -0.069 min
Response: 633619
Conc: 0.27 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T.: 5.951 min
Response: 0
Conc: N.D.




#19 Endosulfan Sulfate

R.T.: 6.962 min
Delta R.T.: -0.014 min
Response: 372867
Conc: 0.13 ng/ml




#19 Endosulfan Sulfate

R.T.: 0.000 min
Exp R.T. : 6.175 min
Response: 0
Conc: N.D.




#20 Methoxychlor

R.T.: 7.398 min
Delta R.T.: 0.062 min
Response: 6178826
Conc: 4.90 ng/ml




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 6.470 min
Response: 0
Conc: N.D.




#21 Endrin ketone

R.T.: 7.514 min
Delta R.T.: 0.056 min
Response: 1163494
Conc: 0.40 ng/ml




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T.: 6.677 min
Response: 0
Conc: N.D.




#22 Toxaphene-1

R.T.: 6.012 min
Delta R.T.: -0.042 min
Response: 282866
Conc: 12.54 ng/ml




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T.: 5.168 min
Response: 0
Conc: N.D.




#24 Toxaphene-3

R.T.: 6.962 min
Delta R.T.: -0.003 min
Response: 372867
Conc: 5.94 ng/ml




#24 Toxaphene-3

R.T.: 0.000 min
Exp R.T. : 6.567 min
Response: 0
Conc: N.D.




#25 Toxaphene-4

R.T.: 7.132 min
Delta R.T.: 0.068 min
Response: 95678
Conc: 2.55 ng/ml




#25 Toxaphene-4

R.T.: 0.000 min
Exp R.T. : 6.881 min
Response: 0
Conc: N.D.




#26 Toxaphene-5

R.T.: 7.722 min
Delta R.T.: -0.029 min
Response: 221044
Conc: 5.01 ng/ml




#26 Toxaphene-5

R.T.: 0.000 min
Exp R.T.: 7.016 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 8.858 min
Delta R.T.: 0.004 min
Response: 55890851
Conc: 22.43 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.773 min
Delta R.T.: 0.001 min
Response: 26019736
Conc: 22.67 ng/ml