

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102820\
 Data File : PL062643.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2020 13:28
 Operator : DD\AJ
 Sample : WATER IDOC 04
 Misc : IDOC FOR ABDUL RAUF(PB132650)
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 WATER IDOC 04

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 06:29:02 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.412	3.951	31749045	57754065	18.458	19.653
2)	SA Decachlor...	8.128	8.991	35753239	49044542	22.286	23.433
Target Compounds							
2)	A alpha-BHC	3.844	4.338	121.8E6	232.7E6	47.759	52.381
3)	MA gamma-BHC...	4.124	4.620	109.2E6	215.1E6	47.135	51.520
4)	MA Heptachlor	4.420	5.126	100.9E6	201.2E6	47.002	50.449
5)	MB Aldrin	4.665	5.429	91623211	189.5E6	48.002	52.368
6)	B beta-BHC	4.373	4.792	45637093	92715781	47.778	49.238
7)	B delta-BHC	4.574	5.006	104.7E6	198.8E6	48.199	56.799
8)	B Heptachlo...	5.104	5.814	89389801	175.4E6	48.992	54.164
9)	A Endosulfan I	5.439	6.169	82156670	159.0E6	49.148	50.410
10)	B gamma-Chl...	5.329	6.048	87284718	170.3E6	48.449	50.575
11)	B alpha-Chl...	5.387	6.120	87081760	171.7E6	48.171	50.962
12)	B 4,4'-DDE	5.552	6.279	80719192	164.6E6	48.789	51.080
13)	MA Dieldrin	5.680	6.426	86079556	162.9E6	48.716	50.900
14)	MA Endrin	5.934	6.643	74468436	132.1E6	44.759	47.323
15)	B Endosulfa...	6.207	6.853	80127180	142.8E6	48.602	49.969
16)	A 4,4'-DDD	6.064	6.769	74070768	144.8E6	52.366	56.569
17)	MA 4,4'-DDT	6.303	7.067	58860509	107.4E6	44.105	47.815
18)	B Endrin al...	6.374	6.977	70857402	122.1E6	48.797	53.535
19)	B Endosulfa...	6.586	7.200	80389691	131.5E6	48.714	54.820
20)	A Methoxychlor	6.847	7.526	36809606	59621226	42.738	48.257
21)	B Endrin ke...	7.073	7.669	96647377	141.4E6	53.603	59.665
22)	Mirex	7.259	8.122	71015008	96340585	50.092	56.535

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102820\
Data File : PL062643.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2020 13:28
Operator : DD\AJ
Sample : WATER IDOC 04
Misc : IDOC FOR ABDUL RAUF(PB132650)
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
WATER IDOC 04

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 06:29:02 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

