

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102820\
 Data File : PL062678.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2020 22:07
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 10/29/2020 12:02:47 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 06:38:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.412	3.952	82602139	148.6E6	48.022	50.562
28)	SA Decachlor...	8.127	8.991	86242675	119.0E6	53.758	56.851
Target Compounds							
2)	A alpha-BHC	3.844	4.338	123.1E6	229.8E6	48.288	51.726
3)	MA gamma-BHC...	4.123	4.621	110.3E6	212.9E6	47.607	50.987
4)	MA Heptachlor	4.420	5.126	100.6E6	195.8E6	46.864	49.106
5)	MB Aldrin	4.665	5.430	92472999	186.5E6	48.447	51.535
6)	B beta-BHC	4.373	4.792	46584646	91431489	48.770	48.556
7)	B delta-BHC	4.574	5.007	105.2E6	195.6E6	48.421	55.889
8)	B Heptachlo...	5.102	5.815	89358647	177.1E6	48.975m	54.694
9)	A Endosulfan I	5.438	6.171	82231743	155.9E6	49.193	49.433
10)	B gamma-Chl...	5.329	6.049	87979073	169.7E6	48.835	50.408
11)	B alpha-Chl...	5.386	6.121	87719375	168.0E6	48.524	49.877
12)	B 4,4'-DDE	5.551	6.280	81045743	160.2E6	48.987	49.721
13)	MA Dieldrin	5.680	6.427	86605097	159.2E6	49.013	49.746
14)	MA Endrin	5.934	6.644	72698144	122.3E6	43.695	43.830
15)	B Endosulfa...	6.207	6.854	80158070	136.1E6	48.621	47.619
16)	A 4,4'-DDD	6.063	6.769	76708561	137.1E6	54.230	53.544
17)	MA 4,4'-DDT	6.303	7.068	54572308	98002094	40.891	43.648
18)	B Endrin al...	6.374	6.977	69659858	116.2E6	47.972	50.945
19)	B Endosulfa...	6.586	7.200	79703075	127.2E6	48.298	53.019
20)	A Methoxychlor	6.846	7.527	35037303	55553905	40.680	44.964
21)	B Endrin ke...	7.072	7.669	95955880	139.5E6	53.220	58.858
22)	Mirex	7.258	8.123	69387707	94275000	48.944	55.323

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102820\
Data File : PL062678.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2020 22:07
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
10/29/2020 12:02:47 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 06:38:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

