

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102822\
 Data File : PL078614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2022 00:24
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 05:31:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.322	2.571	93999485	116.7E6	49.291	45.726
28)	SA Decachlor...	8.772	7.641	74398360	86603386	49.455	45.723
Target Compounds							
2)	A alpha-BHC	3.783	3.058	146.8E6	182.3E6	51.129	46.154
3)	MA gamma-BHC...	4.115	3.379	136.4E6	167.1E6	51.038	46.505
4)	MA Heptachlor	4.686	3.707	135.3E6	158.5E6	52.105	45.713
5)	MB Aldrin	5.024	3.978	127.9E6	162.6E6	51.651	46.438
6)	B beta-BHC	4.330	3.679	58165277	65621248	50.492	45.734
7)	B delta-BHC	4.571	3.899	125.1E6	166.3E6	54.940	45.671
8)	B Heptachlo...	5.458	4.475	118.1E6	139.0E6	55.605	46.141
9)	A Endosulfan I	5.842	4.838	106.7E6	130.5E6	49.990	45.570
10)	B gamma-Chl...	5.714	4.723	116.2E6	138.9E6	51.486	45.326
11)	B alpha-Chl...	5.790	4.786	114.8E6	135.8E6	49.612	45.556
12)	B 4,4'-DDE	5.973	4.980	100.1E6	126.8E6	50.019	46.564
13)	MA Dieldrin	6.119	5.100	107.7E6	137.7E6	49.943	46.471
14)	MA Endrin	6.350	5.371	91659642	119.9E6	48.297	46.025
15)	B Endosulfa...	6.580	5.667	88000992	109.7E6	46.905	48.174
16)	A 4,4'-DDD	6.498	5.531	82274620	106.4E6	49.354	48.458
17)	MA 4,4'-DDT	6.807	5.778	84376089	103.6E6	47.076	47.747
18)	B Endrin al...	6.711	5.846	70975758	86574031	48.779	46.129
19)	B Endosulfa...	6.944	6.068	89051482	111.1E6	48.399	46.687
20)	A Methoxychlor	7.292	6.355	44605014	54172087	48.437	47.737
21)	B Endrin ke...	7.431	6.564	92841866	125.5E6	49.347	48.542
22)	Mirex	7.887	6.737	71415066	89815066	48.921	45.639

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102822\
Data File : PL078614.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2022 00:24
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 05:31:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

