

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102824\
 Data File : PL092691.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2024 00:32
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 01:27:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 28 18:58:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.779	118.3E6	135.0E6	48.290	49.622
28)	SA Decachlor...	9.053	7.915	91426119	137.4E6	47.505	50.350
Target Compounds							
2)	A alpha-BHC	3.996	3.281	163.6E6	201.6E6	48.224	50.527
3)	MA gamma-BHC...	4.328	3.611	156.5E6	194.8E6	48.004	50.462
4)	MA Heptachlor	4.917	3.950	141.6E6	186.9E6	47.148	49.519
5)	MB Aldrin	5.258	4.230	141.7E6	184.2E6	47.211	50.329
6)	B beta-BHC	4.526	3.911	69332252	81489322	47.965	49.500
7)	B delta-BHC	4.773	4.140	149.9E6	194.1E6	47.883	50.516
8)	B Heptachlo...	5.684	4.732	132.6E6	167.6E6	48.004	50.155
9)	A Endosulfan I	6.070	5.102	117.2E6	152.2E6	46.839	49.845
10)	B gamma-Chl...	5.940	4.982	124.6E6	167.7E6	46.833	49.884
11)	B alpha-Chl...	6.019	5.046	125.0E6	165.3E6	47.191	49.705
12)	B 4,4'-DDE	6.192	5.235	112.1E6	162.7E6	47.302	50.491
13)	MA Dieldrin	6.345	5.366	123.9E6	167.8E6	47.003	50.225
14)	MA Endrin	6.574	5.642	104.9E6	146.0E6	46.057	50.459
15)	B Endosulfa...	6.794	5.936	111.4E6	141.7E6	46.735	50.017
16)	A 4,4'-DDD	6.709	5.789	94088352	127.4E6	49.016	51.178
17)	MA 4,4'-DDT	7.023	6.040	96267567	129.2E6	46.512	48.159
18)	B Endrin al...	6.924	6.115	89437162	112.5E6	47.624	48.760
19)	B Endosulfa...	7.158	6.338	102.1E6	132.7E6	47.062	49.188
20)	A Methoxychlor	7.499	6.614	53710650	69401251	47.115	48.599
21)	B Endrin ke...	7.643	6.844	115.4E6	153.3E6	47.628	49.949
22)	Mirex	8.116	7.024	92932289	126.1E6	46.895	48.349
24)	Chlordane-2	5.258f	4.368	141.7E6	3105532	1236.352	25.074 #
25)	Chlordane-3	5.940	4.982	124.6E6	167.7E6	319.242	462.376 #
26)	Chlordane-4	6.019	5.046	125.0E6	165.3E6	261.439	472.595 #
27)	Chlordane-5	0.000	5.936	0	141.7E6	N.D.	1117.525 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102824\
Data File : PL092691.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2024 00:32
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 01:27:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
Quant Title : GC Extractables
QLast Update : Mon Oct 28 18:58:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.540 min
Delta R.T.: 0.000 min
Response: 118322452
Conc: 48.29 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.779 min
Delta R.T.: 0.000 min
Response: 135029150
Conc: 49.62 ng/ml




#2 alpha-BHC

R.T.: 3.996 min
Delta R.T.: 0.000 min
Response: 163554494
Conc: 48.22 ng/ml



#2 alpha-BHC


R.T.: 3.281 min
Delta R.T.: 0.000 min
Response: 201612982
Conc: 50.53 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.328 min
Delta R.T.: 0.000 min
Response: 156472707
Conc: 48.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.611 min
Delta R.T.: 0.000 min
Response: 194753815
Conc: 50.46 ng/ml




#4 Heptachlor

R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 141592348
Conc: 47.15 ng/ml



#4 Heptachlor


R.T.: 3.950 min
Delta R.T.: 0.000 min
Response: 186882032
Conc: 49.52 ng/ml



#5 Aldrin


R.T.: 5.258 min
Delta R.T.: 0.000 min
Response: 141703025
Conc: 47.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.230 min
Delta R.T.: 0.000 min
Response: 184213782
Conc: 50.33 ng/ml




#6 beta-BHC

R.T.: 4.526 min
Delta R.T.: 0.000 min
Response: 69332252
Conc: 47.97 ng/ml



#6 beta-BHC


R.T.: 3.911 min
Delta R.T.: 0.000 min
Response: 81489322
Conc: 49.50 ng/ml



#7 delta-BHC


R.T.: 4.773 min
Delta R.T.: 0.000 min
Response: 149867564
Conc: 47.88 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.140 min
Delta R.T.: 0.000 min
Response: 194140544
Conc: 50.52 ng/ml




#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 132558229
Conc: 48.00 ng/ml



#8 Heptachlor epoxide


R.T.: 4.732 min
Delta R.T.: 0.000 min
Response: 167586934
Conc: 50.15 ng/ml



#9 Endosulfan I


R.T.: 6.070 min
Delta R.T.: 0.000 min
Response: 117161841
Conc: 46.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.102 min
Delta R.T.: 0.000 min
Response: 152182587
Conc: 49.84 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 124605153
Conc: 46.83 ng/ml



#10 gamma-Chlordane


R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 167695960
Conc: 49.88 ng/ml



#11 alpha-Chlordane


R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 124969118
Conc: 47.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.046 min
Delta R.T.: 0.000 min
Response: 165320201
Conc: 49.71 ng/ml




#12 4,4'-DDE

R.T.: 6.192 min
Delta R.T.: 0.000 min
Response: 112066734
Conc: 47.30 ng/ml



#12 4,4'-DDE


R.T.: 5.235 min
Delta R.T.: 0.000 min
Response: 162663129
Conc: 50.49 ng/ml



#13 Dieldrin


R.T.: 6.345 min
Delta R.T.: 0.000 min
Response: 123874204
Conc: 47.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 167765468
Conc: 50.22 ng/ml




#14 Endrin

R.T.: 6.574 min
Delta R.T.: 0.000 min
Response: 104857216
Conc: 46.06 ng/ml



#14 Endrin


R.T.: 5.642 min
Delta R.T.: 0.000 min
Response: 145977961
Conc: 50.46 ng/ml



#15 Endosulfan II


R.T.: 6.794 min
Delta R.T.: 0.000 min
Response: 111409136
Conc: 46.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.936 min
Delta R.T.: 0.000 min
Response: 141711132
Conc: 50.02 ng/ml




#16 4,4'-DDD

R.T.: 6.709 min
Delta R.T.: 0.000 min
Response: 94088352
Conc: 49.02 ng/ml



#16 4,4'-DDD


R.T.: 5.789 min
Delta R.T.: 0.000 min
Response: 127437335
Conc: 51.18 ng/ml



#17 4,4'-DDT


R.T.: 7.023 min
Delta R.T.: 0.000 min
Response: 96267567
Conc: 46.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.040 min
Delta R.T.: 0.000 min
Response: 129174726
Conc: 48.16 ng/ml




#18 Endrin aldehyde

R.T.: 6.924 min
Delta R.T.: 0.000 min
Response: 89437162
Conc: 47.62 ng/ml



#18 Endrin aldehyde


R.T.: 6.115 min
Delta R.T.: 0.000 min
Response: 112504049
Conc: 48.76 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.158 min
Delta R.T.: 0.000 min
Response: 102141500
Conc: 47.06 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.338 min
Delta R.T.: 0.000 min
Response: 132655134
Conc: 49.19 ng/ml




#20 Methoxychlor

R.T.: 7.499 min
Delta R.T.: 0.000 min
Response: 53710650
Conc: 47.11 ng/ml



#20 Methoxychlor


R.T.: 6.614 min
Delta R.T.: 0.000 min
Response: 69401251
Conc: 48.60 ng/ml



#21 Endrin ketone


R.T.: 7.643 min
Delta R.T.: 0.000 min
Response: 115394601
Conc: 47.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.844 min
Delta R.T.: 0.000 min
Response: 153284036
Conc: 49.95 ng/ml




#22 Mirex

R.T.: 8.116 min
Delta R.T.: 0.000 min
Response: 92932289
Conc: 46.90 ng/ml



#22 Mirex


R.T.: 7.024 min
Delta R.T.: 0.000 min
Response: 126145527
Conc: 48.35 ng/ml



#24 Chlordane-2


R.T.: 5.258 min
Delta R.T.: 0.027 min
Response: 141703025
Conc: 1236.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 4.368 min
Delta R.T.: 0.015 min
Response: 3105532
Conc: 25.07 ng/ml




#25 Chlordane-3

R.T.: 5.940 min
Delta R.T.: -0.001 min
Response: 124605153
Conc: 319.24 ng/ml



#25 Chlordane-3


R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 167695960
Conc: 462.38 ng/ml



#26 Chlordane-4


R.T.: 6.019 min
Delta R.T.: -0.004 min
Response: 124969118
Conc: 261.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.046 min
Delta R.T.: 0.001 min
Response: 165320201
Conc: 472.60 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.872 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.936 min
Delta R.T.: -0.005 min
Response: 141711132
Conc: 1117.53 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: -0.001 min
Response: 91426119
Conc: 47.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.915 min
Delta R.T.: 0.000 min
Response: 137418630
Conc: 50.35 ng/ml