

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102918\
 Data File : PL041457.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2018 19:09
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 02:34:57 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102918.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 30 01:26:00 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.474	4.104	317.7E6	152.7E6	49.872	49.577
28)	SA Decachlor...	8.218	9.237	323.4E6	126.2E6	48.825	49.252
Target Compounds							
2)	A alpha-BHC	3.915	4.500	496.9E6	231.9E6	51.149	50.567
3)	MA gamma-BHC...	4.198	4.788	457.1E6	212.3E6	51.168	50.074
4)	MA Heptachlor	4.495	5.304	465.0E6	205.1E6	50.740	50.430
5)	MB Aldrin	4.741	5.613	434.1E6	191.8E6	51.416	50.220
6)	B beta-BHC	4.451	4.957	186.6E6	87294147	49.515	50.827
7)	B delta-BHC	4.653	5.178	427.1E6	203.3E6	51.419	50.340
8)	B Heptachlo...	5.185	6.001	406.7E6	177.4E6	50.374	50.467
9)	A Endosulfan I	5.523	6.362	373.9E6	154.9E6	50.250	49.841
10)	B gamma-Chl...	5.411	6.236	409.7E6	172.2E6	50.314	51.011
11)	B alpha-Chl...	5.468	6.310	404.5E6	167.0E6	49.961	49.333
12)	B 4,4'-DDE	5.635	6.464	366.1E6	162.8E6	51.144	49.739
13)	MA Dieldrin	5.766	6.620	404.2E6	162.6E6	50.827	50.451
14)	MA Endrin	6.023	6.840	369.3E6	135.3E6	50.491	49.787
15)	B Endosulfa...	6.297	7.048	349.2E6	136.6E6	50.523	50.508
16)	A 4,4'-DDD	6.151	6.956	296.6E6	124.5E6	51.039	49.983
17)	MA 4,4'-DDT	6.389	7.256	314.4E6	128.6E6	51.681	51.146
18)	B Endrin al...	6.465	7.172	276.8E6	112.9E6	50.252	51.577
19)	B Endosulfa...	6.678	7.395	316.7E6	123.5E6	49.590	49.770
20)	A Methoxychlor	6.937	7.712	168.7E6	65673233	50.185	50.693
21)	B Endrin ke...	7.168	7.870	365.7E6	133.9E6	50.664	50.876



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102918\
Data File : PL041457.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2018 19:09
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 02:34:57 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102918.M
Quant Title : GC Extractables
QLast Update : Tue Oct 30 01:26:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.474 min
Delta R.T.: 0.000 min
Response: 317696230
Conc: 49.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 4.104 min
Delta R.T.: 0.000 min
Response: 152707316
Conc: 49.58 ng/ml




#2 alpha-BHC

R.T.: 3.915 min
Delta R.T.: 0.000 min
Response: 496933097
Conc: 51.15 ng/ml



#2 alpha-BHC


R.T.: 4.500 min
Delta R.T.: 0.000 min
Response: 231941453
Conc: 50.57 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.198 min
Delta R.T.: 0.000 min
Response: 457132499
Conc: 51.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 4.788 min
Delta R.T.: 0.000 min
Response: 212295995
Conc: 50.07 ng/ml




#4 Heptachlor

R.T.: 4.495 min
Delta R.T.: 0.000 min
Response: 465034982
Conc: 50.74 ng/ml



#4 Heptachlor


R.T.: 5.304 min
Delta R.T.: 0.000 min
Response: 205065540
Conc: 50.43 ng/ml



#5 Aldrin


R.T.: 4.741 min
Delta R.T.: 0.000 min
Response: 434066539
Conc: 51.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 191803260
Conc: 50.22 ng/ml




#6 beta-BHC

R.T.: 4.451 min
Delta R.T.: 0.000 min
Response: 186632185
Conc: 49.52 ng/ml



#6 beta-BHC


R.T.: 4.957 min
Delta R.T.: 0.000 min
Response: 87294147
Conc: 50.83 ng/ml



#7 delta-BHC


R.T.: 4.653 min
Delta R.T.: 0.000 min
Response: 427074908
Conc: 51.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 203289394
Conc: 50.34 ng/ml




#8 Heptachlor epoxide

R.T.: 5.185 min
Delta R.T.: 0.000 min
Response: 406740236
Conc: 50.37 ng/ml



#8 Heptachlor epoxide


R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 177421461
Conc: 50.47 ng/ml



#9 Endosulfan I


R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 373936006
Conc: 50.25 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 6.362 min
Delta R.T.: 0.000 min
Response: 154902395
Conc: 49.84 ng/ml




#10 gamma-Chlordane

R.T.: 5.411 min
Delta R.T.: 0.000 min
Response: 409733353
Conc: 50.31 ng/ml



#10 gamma-Chlordane


R.T.: 6.236 min
Delta R.T.: -0.001 min
Response: 172204496
Conc: 51.01 ng/ml



#11 alpha-Chlordane


R.T.: 5.468 min
Delta R.T.: 0.000 min
Response: 404495264
Conc: 49.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 6.310 min
Delta R.T.: -0.001 min
Response: 166965940
Conc: 49.33 ng/ml




#12 4,4'-DDE

R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 366093502
Conc: 51.14 ng/ml



#12 4,4'-DDE


R.T.: 6.464 min
Delta R.T.: 0.000 min
Response: 162767566
Conc: 49.74 ng/ml



#13 Dieldrin


R.T.: 5.766 min
Delta R.T.: 0.000 min
Response: 404181276
Conc: 50.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 6.620 min
Delta R.T.: -0.001 min
Response: 162626330
Conc: 50.45 ng/ml




#14 Endrin

R.T.: 6.023 min
Delta R.T.: 0.000 min
Response: 369298268
Conc: 50.49 ng/ml



#14 Endrin


R.T.: 6.840 min
Delta R.T.: 0.000 min
Response: 135348894
Conc: 49.79 ng/ml



#15 Endosulfan II


R.T.: 6.297 min
Delta R.T.: -0.002 min
Response: 349210152
Conc: 50.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 7.048 min
Delta R.T.: 0.000 min
Response: 136640461
Conc: 50.51 ng/ml




#16 4,4'-DDD

R.T.: 6.151 min
Delta R.T.: -0.001 min
Response: 296586329
Conc: 51.04 ng/ml



#16 4,4'-DDD


R.T.: 6.956 min
Delta R.T.: -0.001 min
Response: 124450308
Conc: 49.98 ng/ml



#17 4,4'-DDT


R.T.: 6.389 min
Delta R.T.: -0.002 min
Response: 314411527
Conc: 51.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 7.256 min
Delta R.T.: -0.001 min
Response: 128581631
Conc: 51.15 ng/ml




#18 Endrin aldehyde

R.T.: 6.465 min
Delta R.T.: -0.002 min
Response: 276775407
Conc: 50.25 ng/ml



#18 Endrin aldehyde


R.T.: 7.172 min
Delta R.T.: -0.002 min
Response: 112894892
Conc: 51.58 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.678 min
Delta R.T.: -0.001 min
Response: 316700051
Conc: 49.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 7.395 min
Delta R.T.: -0.001 min
Response: 123508324
Conc: 49.77 ng/ml




#20 Methoxychlor

R.T.: 6.937 min
Delta R.T.: -0.002 min
Response: 168696166
Conc: 50.18 ng/ml



#20 Methoxychlor


R.T.: 7.712 min
Delta R.T.: -0.001 min
Response: 65673233
Conc: 50.69 ng/ml



#21 Endrin ketone


R.T.: 7.168 min
Delta R.T.: -0.001 min
Response: 365674279
Conc: 50.66 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 7.870 min
Delta R.T.: -0.002 min
Response: 133889915
Conc: 50.88 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.218 min
Delta R.T.: -0.001 min
Response: 323415142
Conc: 48.82 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.237 min
Delta R.T.: -0.001 min
Response: 126191276
Conc: 49.25 ng/ml