

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102920\
 Data File : PL062679.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2020 08:39
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 01:13:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

16) A	4,4'-DDD	6.077	0.000	4665858	0	3.299	N.D.	#
17) MA	4,4'-DDT	6.307	7.045f	5857982	841286	4.389	0.375	#
18) B	Endrin al...	6.389f	0.000	1416458	0	0.975	N.D.	#
19) B	Endosulfa...	0.000	7.180f	0	3459481	N.D.	1.442	#
20) A	Methoxychlor	6.842	0.000	711470	0	0.826	N.D.	#
21) B	Endrin ke...	7.059	7.641f	338865	9847783	0.188	4.156	#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102920\
Data File : PL062679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2020 08:39
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 01:13:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.412 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.952 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 6.077 min
Delta R.T.: 0.014 min
Response: 4665858
Conc: 3.30 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T. : 6.768 min
Response: 0
Conc: N.D.



#17 4,4'-DDT


R.T.: 6.307 min
Delta R.T.: 0.006 min
Response: 5857982
Conc: 4.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
HEXANE




#17 4,4'-DDT

R.T.: 7.045 min
Delta R.T.: -0.021 min
Response: 841286
Conc: 0.37 ng/ml




#18 Endrin aldehyde

R.T.: 6.389 min
Delta R.T.: 0.017 min
Response: 1416458
Conc: 0.98 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T. : 6.976 min
Response: 0
Conc: N.D.



#19 Endosulfan Sulfate


R.T.: 0.000 min
Exp R.T. : 6.585 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE




#19 Endosulfan Sulfate

R.T.: 7.180 min
Delta R.T.: -0.019 min
Response: 3459481
Conc: 1.44 ng/ml




#20 Methoxychlor

R.T.: 6.842 min
Delta R.T.: -0.003 min
Response: 711470
Conc: 0.83 ng/ml



#20 Methoxychlor


R.T.: 0.000 min
Exp R.T. : 7.526 min
Response: 0
Conc: N.D.



#21 Endrin ketone


R.T.: 7.059 min
Delta R.T.: -0.012 min
Response: 338865
Conc: 0.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
HEXANE




#21 Endrin ketone

R.T.: 7.641 min
Delta R.T.: -0.027 min
Response: 9847783
Conc: 4.16 ng/ml



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.126 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.989 min
Response: 0
Conc: N.D.