

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102921\
 Data File : PL070655.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2021 07:21
 Operator : AR\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 07:42:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102921.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 29 06:36:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	50.000	53.353	-6.7	103	0.00
23	Chlordane-1	500.000	543.336	-8.7	104	0.00
24	Chlordane-2	500.000	524.274	-4.9	102	0.00
25	Chlordane-3	500.000	538.715	-7.7	103	0.00
26	Chlordane-4	500.000	538.017	-7.6	103	0.00
27	Chlordane-5	500.000	498.743	0.3	102	0.00
28 SA	Decachlorobiphenyl	50.000	51.294	-2.6	101	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	50.000	52.948	-5.9	101	0.00
23	Chlordane-1	500.000	543.343	-8.7	103	0.00
24	Chlordane-2	500.000	537.747	-7.5	101	0.00
25	Chlordane-3	500.000	535.494	-7.1	101	0.00
26	Chlordane-4	500.000	534.641	-6.9	101	0.00
27	Chlordane-5	500.000	531.948	-6.4	102	0.00
28 SA	Decachlorobiphenyl	50.000	53.141	-6.3	101	0.00

Evaluate Continuing Calibration Report - Not Found

2 A	alpha-BHC	50.000	0.000	100.0#	0	-5.29#
3 MA	gamma-BHC (Lindane)	50.000	0.000	100.0#	0	-5.71#
4 MA	Heptachlor	50.000	0.000	100.0#	0	-6.12#
5 MB	Aldrin	50.000	0.000	100.0#	0	-6.44#
6 B	beta-BHC	50.000	0.000	100.0#	0	-6.09#
7 B	delta-BHC	50.000	0.000	100.0#	0	-6.35#
8 B	Heptachlor epoxide	50.000	0.000	100.0#	0	-7.02#
9 A	Endosulfan I	50.000	0.000	100.0#	0	-7.42#
10 B	gamma-Chlordane	50.000	0.000	100.0#	0	-7.30#
11 B	alpha-Chlordane	50.000	0.000	100.0#	0	-7.37#
12 B	4,4'-DDE	50.000	0.000	100.0#	0	-7.58#
13 MA	Dieldrin	50.000	0.000	100.0#	0	-7.71#
14 MA	Endrin	50.000	0.000	100.0#	0	-8.00#
15 B	Endosulfan II	50.000	0.000	100.0#	0	-8.31#
16 A	4,4'-DDD	50.000	0.000	100.0#	0	-8.17#
17 MA	4,4'-DDT	50.000	0.000	100.0#	0	-8.43#
18 B	Endrin aldehyde	50.000	0.000	100.0#	0	-8.50#
19 B	Endosulfan Sulfate	50.000	0.000	100.0#	0	-8.73#
20 A	Methoxychlor	50.000	0.000	100.0#	0	-9.02#
21 B	Endrin ketone	50.000	0.000	100.0#	0	-9.25#
22	Mirex	50.000	0.000	100.0#	0	-9.43#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102921\
 Data File : PL070655.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2021 07:21
 Operator : AR\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 07:42:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102921.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 29 06:36:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)

Signal #2					
2 A alpha-BHC	50.000	0.000	100.0#	0	-6.10#
3 MA gamma-BHC (Lindane)	50.000	0.000	100.0#	0	-6.49#
4 MA Heptachlor	50.000	0.000	100.0#	0	-7.14#
5 MB Aldrin	50.000	0.000	100.0#	0	-7.51#
6 B beta-BHC	50.000	0.000	100.0#	0	-6.75#
7 B delta-BHC	50.000	0.000	100.0#	0	-7.02#
8 B Heptachlor epoxide	50.000	0.000	100.0#	0	-7.98#
9 A Endosulfan I	50.000	0.000	100.0#	0	-8.38#
10 B gamma-Chlordane	50.000	0.000	100.0#	0	-8.25#
11 B alpha-Chlordane	50.000	0.000	100.0#	0	-8.33#
12 B 4,4'-DDE	50.000	0.000	100.0#	0	-8.52#
13 MA Dieldrin	50.000	0.000	100.0#	0	-8.67#
14 MA Endrin	50.000	0.000	100.0#	0	-8.91#
15 B Endosulfan II	50.000	0.000	100.0#	0	-9.15#
16 A 4,4'-DDD	50.000	0.000	100.0#	0	-9.06#
17 MA 4,4'-DDT	50.000	0.000	100.0#	0	-9.37#
18 B Endrin aldehyde	50.000	0.000	100.0#	0	-9.28#
19 B Endosulfan Sulfate	50.000	0.000	100.0#	0	-9.52#
20 A Methoxychlor	50.000	0.000	100.0#	0	-9.86#
21 B Endrin ketone	50.000	0.000	100.0#	0	-10.02#
22 Mirex	50.000	0.000	100.0#	0	-10.48#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0