

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102924\
 Data File : PL092694.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2024 10:02
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 12:30:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 28 18:58:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.541	2.779	49453430	52803176	20.183	19.405
28) SA Decachlor...	9.054	7.916	39703848	55817184	20.630	20.452
Target Compounds						
2) A alpha-BHC	3.996	3.281	35290534	36492710	10.405	9.146
3) MA gamma-BHC...	4.329	3.611	33338334	34494865	10.228	8.938
6) B beta-BHC	4.526	3.911	15793927	17190837	10.927	10.442
14) MA Endrin	6.575	5.642	96595274	137.1E6	42.429	47.403
16) A 4,4'-DDD	6.710	5.789	3848508	3505990	2.005	1.408 #
17) MA 4,4'-DDT	7.024	6.040	184.3E6	274.9E6	89.047	102.500
18) B Endrin al...	6.924	6.116	3058704	3670909	1.629	1.591
20) A Methoxychlor	7.500	6.615	238.4E6	331.0E6	209.097	231.819
21) B Endrin ke...	7.643	6.843	6180696	6769458	2.551	2.206

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102924\
Data File : PL092694.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2024 10:02
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 12:30:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
Quant Title : GC Extractables
QLast Update : Mon Oct 28 18:58:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

