

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102925\
 Data File : PL097675.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2025 15:16
 Operator : AR\AJ
 Sample : IDOC-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 IDOC-02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 01:23:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100725.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 07 16:47:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.523	2.821	89238413	129.4E6	21.890	21.253
2)	SA Decachlor...	8.995	7.979	68028939	108.9E6	22.877	20.991
Target Compounds							
2)	A alpha-BHC	3.970	3.325	322.1E6	458.0E6	47.744	46.417
3)	MA gamma-BHC...	4.297	3.657	300.6E6	426.1E6	47.297	46.375
4)	MA Heptachlor	4.888	4.004	303.8E6	413.5E6	47.100	44.987
5)	MB Aldrin	5.227	4.286	286.4E6	398.8E6	47.638	45.989
6)	B beta-BHC	4.485	3.953	120.8E6	177.0E6	46.723	44.501
7)	B delta-BHC	4.731	4.185	284.6E6	416.6E6	46.868	46.119
8)	B Heptachlo...	5.647	4.788	263.7E6	357.8E6	45.290	45.297
9)	A Endosulfan I	6.029	5.159	235.6E6	337.1E6	47.384	45.713
10)	B gamma-Chl...	5.901	5.040	258.6E6	388.0E6	47.642	44.637
11)	B alpha-Chl...	5.981	5.104	256.2E6	374.3E6	47.205	44.954
12)	B 4,4'-DDE	6.151	5.292	218.1E6	347.8E6	46.823	45.811
13)	MA Dieldrin	6.300	5.423	252.3E6	365.4E6	47.953	44.197
14)	MA Endrin	6.527	5.698	201.4E6	311.3E6	45.018	42.064
15)	B Endosulfa...	6.741	5.990	220.4E6	309.4E6	45.049	44.810
16)	A 4,4'-DDD	6.661	5.845	180.5E6	292.4E6	47.823	45.836
17)	MA 4,4'-DDT	6.975	6.097	185.8E6	289.3E6	45.824	43.410
18)	B Endrin al...	6.870	6.168	157.1E6	234.6E6	47.948	40.748
19)	B Endosulfa...	7.103	6.391	187.8E6	291.2E6	46.844	43.739
20)	A Methoxychlor	7.448	6.669	91094848	149.9E6	43.607	41.132
21)	B Endrin ke...	7.583	6.895	205.6E6	332.0E6	48.027	42.031
22)	Mirex	8.060	7.085	139.2E6	222.0E6	44.809	41.504

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102925\
Data File : PL097675.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2025 15:16
Operator : AR\AJ
Sample : IDOC-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
IDOC-02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 01:23:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100725.M
Quant Title : GC Extractables
QLast Update : Tue Oct 07 16:47:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

