

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL103019\
 Data File : PL053906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2019 09:40
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 10/31/2019 10:21:38 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 13:15:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	4.044	577.9E6	130.0E6	53.316	51.936
28)	SA Decachlor...	8.324	9.120	469.8E6	137.8E6	47.878	48.930
Target Compounds							
2)	A alpha-BHC	3.985	4.435	945.9E6	184.3E6	54.247	53.298
3)	MA gamma-BHC...	4.272	4.720	835.4E6	170.6E6	53.645	52.976
4)	MA Heptachlor	4.578	5.230	827.0E6	160.4E6	52.398	49.784
5)	MB Aldrin	4.830	5.535	795.6E6	154.6E6	55.654	51.847
6)	B beta-BHC	4.522	4.891	340.4E6	78894363	51.651	51.077
7)	B delta-BHC	4.730	5.108	806.5E6	165.0E6	52.986	54.617
8)	B Heptachlor...	5.275	5.921	725.1E6	156.0E6	53.717	53.482
9)	A Endosulfan I	5.617	6.278	654.8E6	138.4E6	53.272	51.049
10)	B gamma-Chl...	5.503	6.154	718.4E6	150.3E6	52.658	51.527m
11)	B alpha-Chl...	5.562	6.228	692.6E6	149.2E6	52.061	51.102
12)	B 4,4'-DDE	5.725	6.384	627.6E6	143.2E6	52.816	50.861
13)	MA Dieldrin	5.861	6.535	681.9E6	146.9E6	52.322	51.486
14)	MA Endrin	6.120	6.753	575.9E6	117.8E6	49.682	48.525
15)	B Endosulfa...	6.393	6.962	562.3E6	132.1E6	50.292	51.697
16)	A 4,4'-DDD	6.243	6.874	520.6E6	120.5E6	58.500	57.816
17)	MA 4,4'-DDT	6.484	7.172	441.3E6	102.6E6	44.550	44.387
18)	B Endrin al...	6.560	7.086	449.9E6	110.2E6	49.066	50.169
19)	B Endosulfa...	6.775	7.309	505.8E6	119.8E6	49.726	48.810
20)	A Methoxychlor	7.030	7.630	223.3E6	57084028	45.284	43.948
21)	B Endrin ke...	7.268	7.780	556.8E6	133.5E6	51.720	51.790
22)	Mirex	7.461	8.234	411.3E6	101.9E6	48.083	50.385

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL103019\
Data File : PL053906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2019 09:40
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
10/31/2019 10:21:38 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 13:15:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

