

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : AR1268PP110120.M
 Title : GC EXTRACTABLES
 Last Update : Sat Oct 31 02:37:11 2020
 Response Via : Initial Calibration

Calibration Files

500 =PP030997.D 1000 =PP030995.D 750 =PP030996.D
 250 =PP030998.D 50 =PP030999.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	5.793	5.471	5.557	5.981	6.994	5.959 E4	10.27
2) SA	Decachlorobiphenyl	5.593	5.295	5.399	5.763	6.642	5.739 E4	9.35
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	4.679	4.451	4.502	4.760	5.964	4.871 E3	12.81
42) L9	AR-1268-2	4.737	4.518	4.473	4.666	5.824	4.844 E3	11.53
43) L9	AR-1268-3	3.931	3.721	3.787	3.948	4.845	4.046 E3	11.28
44) L9	AR-1268-4	1.595	1.525	1.528	1.606	1.777	1.606 E3	6.38
45) L9	AR-1268-5	1.260	1.236	1.246	1.292	1.392	1.285 E4	4.95

Signal #2 Calibration Files

500 =PP030997.D 1000 =PP030995.D 750 =PP030996.D
 250 =PP030998.D 50 =PP030999.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	5.188	4.927	5.003	5.211	5.990	5.264 E4	8.05
2) SA	Decachlorobiphenyl	6.130	5.949	5.994	6.222	7.273	6.314 E4	8.67
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : AR1268PP110120.M
 Title : GC EXTRACTABLES
 Last Update : Sat Oct 31 02:37:11 2020
 Response Via : Initial Calibration

Calibration Files

500 =PP030997.D 1000 =PP030995.D 750 =PP030996.D
 250 =PP030998.D 50 =PP030999.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	5.194	5.019	5.030	5.263	6.391	5.379 E3	10.69
42) L9	AR-1268-2	5.153	5.021	4.995	5.103	6.122	5.279 E3	9.01
43) L9	AR-1268-3	4.289	4.137	4.162	4.260	5.200	4.410 E3	10.13
44) L9	AR-1268-4	1.773	1.738	1.746	1.781	1.955	1.799 E3	4.96
45) L9	AR-1268-5	1.418	1.394	1.399	1.408	1.526	1.429 E4	3.84

(#) = Out of Range