

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_L\Data\PL103118\
 Data File : PL041587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2018 00:29
 Operator : AJ\SJ
 Sample : WATER IDOC 02
 Misc : SOMINA IDOC 2
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 WATER IDOC 02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 01:33:58 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102918.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 30 01:26:00 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.473	4.104	110.0E6	61352273	17.265	19.918
28)	SA Decachlor...	8.217	9.235	118.7E6	46913406	17.916	18.310
Target Compounds							
2)	A alpha-BHC	3.914	4.499	447.7E6	238.3E6	46.079	51.957
3)	MA gamma-BHC...	4.197	4.788	412.8E6	219.4E6	46.203	51.744
4)	MA Heptachlor	4.494	5.303	420.5E6	212.8E6	45.881	52.333
5)	MB Aldrin	4.740	5.613	407.2E6	209.5E6	48.238	54.846
6)	B beta-BHC	4.450	4.956	171.0E6	88296043	45.355	51.410
7)	B delta-BHC	4.653	5.176	373.0E6	204.1E6	44.903	50.530
8)	B Heptachlor...	5.184	6.000	372.8E6	179.1E6	46.166	50.957
9)	A Endosulfan I	5.522	6.360	341.7E6	167.3E6	45.923	53.844
10)	B gamma-Chl...	5.410	6.235	372.7E6	181.2E6	45.766	53.681
11)	B alpha-Chl...	5.467	6.309	365.5E6	177.0E6	45.151	52.286
12)	B 4,4'-DDE	5.634	6.463	334.2E6	176.4E6	46.688	53.916
13)	MA Dieldrin	5.765	6.619	367.6E6	177.7E6	46.229	55.122
14)	MA Endrin	6.022	6.839	317.8E6	139.4E6	43.456	51.286
15)	B Endosulfa...	6.296	7.047	310.7E6	140.5E6	44.952	51.918
16)	A 4,4'-DDD	6.150	6.955	277.3E6	141.7E6	47.720	56.926
17)	MA 4,4'-DDT	6.388	7.255	270.1E6	128.2E6	44.391	50.983
18)	B Endrin al...	6.464	7.171	251.9E6	116.2E6	45.739	53.085
19)	B Endosulfa...	6.677	7.394	306.1E6	134.1E6	47.937	54.052
20)	A Methoxychlor	6.936	7.711	144.1E6	62699764	42.860	48.398
21)	B Endrin ke...	7.167	7.868	330.6E6	141.6E6	45.812	53.791

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_L\Data\PL103118\
Data File : PL041587.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2018 00:29
Operator : AJ\SJ
Sample : WATER IDOC 02
Misc : SOMINA IDOC 2
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
WATER IDOC 02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 01:33:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102918.M
Quant Title : GC Extractables
QLast Update : Tue Oct 30 01:26:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

