

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL103122\
 Data File : PL078661.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2022 18:05
 Operator : AR\AJ
 Sample : N5310-17MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 BIN 1-2-(2840)-COMP-1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 22:14:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.323	2.569	36324479	53147949	19.048	20.816
28)	SA Decachlor...	8.776	7.639	30777737	42967950	20.459	22.685
Target Compounds							
2)	A alpha-BHC	3.783	3.056	140.7E6	198.5E6	49.015	50.264
3)	MA gamma-BHC...	4.116	3.377	129.0E6	182.1E6	48.283	50.673
4)	MA Heptachlor	4.688	3.704	129.6E6	173.9E6	49.908	50.136
5)	MB Aldrin	5.025	3.976	126.3E6	186.3E6	51.006	53.203
6)	B beta-BHC	4.331	3.677	54312340	68691483	47.147	47.874
7)	B delta-BHC	4.572	3.897	110.0E6	183.8E6	48.320	50.470
8)	B Heptachlo...	5.459	4.473	114.1E6	155.3E6	53.749	51.530
9)	A Endosulfan I	5.843	4.836	103.4E6	141.8E6	48.471	49.520
10)	B gamma-Chl...	5.715	4.720	116.6E6	157.8E6	51.651	51.482
11)	B alpha-Chl...	5.791	4.783	113.0E6	156.2E6	48.831	52.398
12)	B 4,4'-DDE	5.974	4.978	97645497	143.3E6	48.790	52.619
13)	MA Dieldrin	6.121	5.097	109.5E6	160.8E6	50.782	54.277
14)	MA Endrin	6.352	5.368	96698166	144.8E6	50.952	55.589
15)	B Endosulfa...	6.582	5.664	91036060	132.2E6	48.522	58.059
16)	A 4,4'-DDD	6.499	5.529	84453998	116.7E6	50.661	53.140
17)	MA 4,4'-DDT	6.809	5.775	81509659	117.0E6	45.477	53.932
18)	B Endrin al...	6.713	5.843	75945189	105.8E6	52.194	56.367
19)	B Endosulfa...	6.947	6.066	90762707	129.3E6	49.329	54.335
20)	A Methoxychlor	7.293	6.352	45356417	65069243	49.253	57.340
21)	B Endrin ke...	7.433	6.562	95007982	143.3E6	50.499	55.423
22)	Mirex	7.890	6.734	74626074	106.2E6	51.120	53.948

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL103122\
Data File : PL078661.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2022 18:05
Operator : AR\AJ
Sample : N5310-17MS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
BIN 1-2-(2840)-COMP-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 22:14:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

