

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110119\
 Data File : PL053991.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2019 09:42
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 22:34:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	4.045	588.4E6	132.9E6	54.284	53.121
28)	SA Decachlor...	8.327	9.124	485.2E6	141.5E6	49.455	50.238
Target Compounds							
2)	A alpha-BHC	3.986	4.436	964.8E6	189.4E6	55.331	54.796
3)	MA gamma-BHC...	4.273	4.721	851.9E6	174.4E6	54.699	54.147
4)	MA Heptachlor	4.580	5.231	834.9E6	161.9E6	52.900	50.230
5)	MB Aldrin	4.831	5.537	794.8E6	156.6E6	55.596	52.523
6)	B beta-BHC	4.523	4.891	345.2E6	81147091	52.382	52.535
7)	B delta-BHC	4.732	5.109	814.3E6	167.6E6	53.496	55.477
8)	B Heptachlor...	5.277	5.923	724.3E6	161.5E6	53.651	55.379
9)	A Endosulfan I	5.619	6.280	668.3E6	139.5E6	54.368	51.442
10)	B gamma-Chl...	5.505	6.157	728.1E6	149.6E6	53.367	51.288
11)	B alpha-Chl...	5.564	6.230	697.5E6	150.3E6	52.426	51.494
12)	B 4,4'-DDE	5.727	6.386	622.4E6	144.4E6	52.377	51.285
13)	MA Dieldrin	5.863	6.537	687.6E6	147.8E6	52.756	51.795
14)	MA Endrin	6.122	6.755	573.9E6	117.7E6	49.512	48.490
15)	B Endosulfa...	6.394	6.964	563.1E6	132.8E6	50.369	51.977
16)	A 4,4'-DDD	6.244	6.876	527.5E6	122.2E6	59.270	58.650
17)	MA 4,4'-DDT	6.486	7.174	427.8E6	100.0E6	43.180	43.248
18)	B Endrin al...	6.563	7.088	450.9E6	110.2E6	49.177	50.147
19)	B Endosulfa...	6.776	7.310	505.3E6	122.4E6	49.675	49.903
20)	A Methoxychlor	7.032	7.632	215.2E6	55141708	43.645	42.452
21)	B Endrin ke...	7.270	7.782	559.2E6	137.0E6	51.945	53.128
22)	Mirex	7.463	8.236	408.9E6	105.2E6	47.805	51.994

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110119\
Data File : PL053991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2019 09:42
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 22:34:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

