

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110122\
 Data File : PL078711.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2022 18:51
 Operator : AR\AJ
 Sample : N5380-04
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-D

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 22:30:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.322	2.569	23465888	34727750	12.305	13.602
28)	SA Decachlor...	8.770	7.637	18821892	28582407	12.512	15.090
Target Compounds							
11)	B alpha-Chl...	5.791	4.778	1597758	1419075	0.691	0.476 #
12)	B 4,4'-DDE	5.971	4.976	698278	994139	0.349	0.365



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110122\
Data File : PL078711.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2022 18:51
Operator : AR\AJ
Sample : N5380-04
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-D

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 22:30:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.322 min
Delta R.T.: -0.001 min
Response: 23465888
Conc: 12.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-D




#1 Tetrachloro-m-xylene

R.T.: 2.569 min
Delta R.T.: -0.005 min
Response: 34727750
Conc: 13.60 ng/ml




#11 alpha-Chlordane

R.T.: 5.791 min
Delta R.T.: 0.000 min
Response: 1597758
Conc: 0.69 ng/ml



#11 alpha-Chlordane


R.T.: 4.778 min
Delta R.T.: -0.009 min
Response: 1419075
Conc: 0.48 ng/ml



#12 4,4'-DDE


R.T.: 5.971 min
Delta R.T.: -0.003 min
Response: 698278
Conc: 0.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-D




#12 4,4'-DDE

R.T.: 4.976 min
Delta R.T.: -0.006 min
Response: 994139
Conc: 0.37 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.770 min
Delta R.T.: -0.007 min
Response: 18821892
Conc: 12.51 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.637 min
Delta R.T.: -0.007 min
Response: 28582407
Conc: 15.09 ng/ml