

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX110123.M
 Title : GC Extractables
 Last Update : Thu Nov 02 05:41:57 2023
 Response Via : Initial Calibration

Calibration Files

500 =PL086355.D 1000 =PL086353.D 250 =PL086356.D
 750 =PL086354.D 100 =PL086357.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	3.110	3.150	3.202	3.094	3.220	3.155	E6	1.75
2)	Toxaphene-1	2.477	2.517	2.623	2.465	2.757	2.568	E4	4.78
3)	Toxaphene-2	2.019	2.086	2.046	2.071	2.135	2.072	E4	2.12
4)	Toxaphene-3	6.537	6.776	6.660	6.666	6.753	6.678	E4	1.41
5)	Toxaphene-4	3.959	4.146	4.019	4.064	4.133	4.064	E4	1.92
6)	Toxaphene-5	4.626	4.781	4.652	4.667	4.501	4.645	E4	2.16
7) SA	Decachlorobiphenyl	2.649	2.566	2.801	2.582	2.757	2.671	E6	3.92

Signal #2 Calibration Files

500 =PL086355.D 1000 =PL086353.D 250 =PL086356.D
 750 =PL086354.D 100 =PL086357.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	9.335	9.525	9.287	9.380	9.115	9.328	E5	1.60
2)	Toxaphene-1	5.774	6.020	5.735	5.837	5.848	5.843	E3	1.87
3)	Toxaphene-2	6.851	7.307	6.804	7.241	6.506	6.942	E3	4.78
4)	Toxaphene-3	2.643	2.743	2.429	2.695	2.243	2.551	E4	8.23
5)	Toxaphene-4	3.686	3.881	3.556	3.673	3.241	3.607	E4	6.54
6)	Toxaphene-5	2.509	2.668	2.397	2.565	2.412	2.510	E4	4.46
7) SA	Decachlorobiphenyl	1.136	1.109	1.200	1.101	1.264	1.162	E6	5.93

(#) = Out of Range