

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110123\
 Data File : PL086342.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2023 08:36
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RESCHK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 05:45:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110123.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 02 05:32:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.353	2.618	58369602	17467459	18.393	18.348
28)	SA Decachlor...	8.855	7.770	51356727	22608832	19.418	19.711
Target Compounds							
9)	A Endosulfan I	5.879	4.929	34306676	12213871	9.333	10.291
10)	B gamma-Chl...	5.753	4.813	38045967	10944139	9.589	9.248
12)	B 4,4'-DDE	6.016	5.076	59585434	18015149	18.775	18.255
13)	MA Dieldrin	6.156	5.194	69684831	21338483	18.124	17.869
19)	B Endosulfa...	6.976	6.173	59123098	20016833	19.146	18.986
20)	A Methoxychlor	7.337	6.468	132.5E6	55111470	95.523	94.315
21)	B Endrin ke...	7.457	6.674	59363923	23117259	18.601	18.536



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110123\
Data File : PL086342.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2023 08:36
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
RESCHK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 05:45:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110123.M
Quant Title : GC Extractables
QLast Update : Thu Nov 02 05:32:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.353 min
Delta R.T.: 0.000 min
Response: 58369602
Conc: 18.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#1 Tetrachloro-m-xylene

R.T.: 2.618 min
Delta R.T.: 0.000 min
Response: 17467459
Conc: 18.35 ng/ml




#9 Endosulfan I

R.T.: 5.879 min
Delta R.T.: 0.000 min
Response: 34306676
Conc: 9.33 ng/ml



#9 Endosulfan I


R.T.: 4.929 min
Delta R.T.: 0.000 min
Response: 12213871
Conc: 10.29 ng/ml



#10 gamma-Chlordane


R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 38045967
Conc: 9.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#10 gamma-Chlordane

R.T.: 4.813 min
Delta R.T.: 0.000 min
Response: 10944139
Conc: 9.25 ng/ml




#12 4,4'-DDE

R.T.: 6.016 min
Delta R.T.: 0.000 min
Response: 59585434
Conc: 18.78 ng/ml



#12 4,4'-DDE


R.T.: 5.076 min
Delta R.T.: 0.000 min
Response: 18015149
Conc: 18.26 ng/ml



#13 Dieldrin


R.T.: 6.156 min
Delta R.T.: 0.000 min
Response: 69684831
Conc: 18.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#13 Dieldrin

R.T.: 5.194 min
Delta R.T.: 0.000 min
Response: 21338483
Conc: 17.87 ng/ml




#19 Endosulfan Sulfate

R.T.: 6.976 min
Delta R.T.: 0.000 min
Response: 59123098
Conc: 19.15 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.173 min
Delta R.T.: 0.000 min
Response: 20016833
Conc: 18.99 ng/ml



#20 Methoxychlor


R.T.: 7.337 min
Delta R.T.: 0.000 min
Response: 132491233
Conc: 95.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#20 Methoxychlor

R.T.: 6.468 min
Delta R.T.: 0.000 min
Response: 55111470
Conc: 94.31 ng/ml




#21 Endrin ketone

R.T.: 7.457 min
Delta R.T.: 0.000 min
Response: 59363923
Conc: 18.60 ng/ml



#21 Endrin ketone


R.T.: 6.674 min
Delta R.T.: 0.000 min
Response: 23117259
Conc: 18.54 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.855 min
Delta R.T.: 0.000 min
Response: 51356727
Conc: 19.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK



#28 Decachlorobiphenyl

R.T.: 7.770 min
Delta R.T.: 0.000 min
Response: 22608832
Conc: 19.71 ng/ml