

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110320\
 Data File : PL062795.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Nov 2020 08:56
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 17:00:57 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.411	3.951	88105716	153.1E6	51.222	52.098
28)	SA Decachlor...	8.127	8.992	86442283	110.9E6	53.882	52.992
Target Compounds							
2)	A alpha-BHC	3.843	4.338	132.3E6	231.6E6	51.874	52.117
3)	MA gamma-BHC...	4.123	4.620	117.6E6	215.3E6	50.771	51.560
4)	MA Heptachlor	4.420	5.126	110.0E6	197.3E6	51.283	49.463
5)	MB Aldrin	4.664	5.429	98298844	183.4E6	51.499	50.677
6)	B beta-BHC	4.372	4.792	49181988	93473602	51.489	49.640
7)	B delta-BHC	4.573	5.007	111.6E6	187.1E6	51.330	53.479
8)	B Heptachlo...	5.103	5.815	96408620	164.8E6	52.839	50.881
9)	A Endosulfan I	5.438	6.171	87611959	152.9E6	52.412	48.467
10)	B gamma-Chl...	5.329	6.049	93075530	165.0E6	51.663	49.010
11)	B alpha-Chl...	5.386	6.121	93170327	164.0E6	51.539	48.666
12)	B 4,4'-DDE	5.551	6.280	86677506	158.8E6	52.391	49.287
13)	MA Dieldrin	5.680	6.427	92128599	156.9E6	52.139	49.010
14)	MA Endrin	5.934	6.644	82430516	129.6E6	49.544	46.443
15)	B Endosulfa...	6.207	6.853	85540522	132.0E6	51.885	46.189
16)	A 4,4'-DDD	6.064	6.770	76180922	126.3E6	53.857	49.321
17)	MA 4,4'-DDT	6.302	7.068	67712608	109.8E6	50.738	48.887
18)	B Endrin al...	6.374	6.978	72949784	111.3E6	50.238	48.802
19)	B Endosulfa...	6.586	7.201	83889273	122.7E6	50.835	51.152
20)	A Methoxychlor	6.846	7.527	43127323	59896219	50.073	48.479
21)	B Endrin ke...	7.072	7.670	98330223	128.0E6	54.537	54.014
22)	Mirex	7.258	8.123	74609728	87748636	52.628	51.493

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110320\
Data File : PL062795.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2020 08:56
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 17:00:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

