

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110422\
 Data File : PL078813.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2022 12:08
 Operator : AR\AJ
 Sample : N5360-19
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 DISSOLVED-20221179-COMP-14-MODIFIED-ELUTRIATE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 03:07:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.332	2.570	25526709	35788427	13.386	14.017
28)	SA Decachlor...	8.783	7.642	7821656	10841616	5.199	5.724

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110422\
Data File : PL078813.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2022 12:08
Operator : AR\AJ
Sample : N5360-19
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
DISSOLVED-20221179-COMP-14-MODIFIED-ELUTRIATE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 03:07:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.332 min
Delta R.T.: 0.009 min
Response: 25526709
Conc: 13.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
DISSOLVED-20221179-COMP-14-MODIFIED-ELUTRIATE




#1 Tetrachloro-m-xylene

R.T.: 2.570 min
Delta R.T.: -0.004 min
Response: 35788427
Conc: 14.02 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.783 min
Delta R.T.: 0.007 min
Response: 7821656
Conc: 5.20 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.642 min
Delta R.T.: -0.002 min
Response: 10841616
Conc: 5.72 ng/ml