

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110423\
 Data File : PL086425.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Nov 2023 13:59
 Operator : AR\AJ
 Sample : INDA101
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 INDA1306

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 16:45:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110423CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Nov 03 16:33:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.351	2.618	13681498	3978121	5.172	4.954
27)	SA Decachlor...	8.851	7.766	25236552	11490703	10.635	10.879
Target Compounds							
2)	A alpha-BHC	3.805	3.115	18237403	4730645	4.671	4.331
3)	MA gamma-BHC...	4.136	3.442	18287273	4845142	4.895	4.502
4)	MA Heptachlor	4.723	3.778	20219125	5684346	5.123	4.891
9)	A Endosulfan I	5.877	4.926	16233123	4863517	5.248	5.000
13)	MA Dieldrin	6.154	5.192	32795675	9777027	9.758	9.072
14)	MA Endrin	6.382	5.465	29723839	9265020	10.162	9.465
16)	A 4,4'-DDD	6.533	5.629	21733667	6694878	9.752	8.771
17)	MA 4,4'-DDT	6.847	5.880	23005939	7869529	9.667	8.927
20)	A Methoxychlor	7.334	6.465	65850662	28035533	51.661	51.489

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110423\
Data File : PL086425.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2023 13:59
Operator : AR\AJ
Sample : INDA101
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 16:45:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110423CLP.M
Quant Title : GC Extractables
QLast Update : Fri Nov 03 16:33:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

