

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : PL110423CLP.M
 Title : GC Extractables
 Last Update : Fri Nov 03 16:24:02 2023
 Response Via : Initial Calibration

Calibration Files

3 =PL086422.D 1 =PL086420.D 2 =PL086421.D
 4 =PL086423.D 5 =PL086424.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.250	3.163	3.118	3.099	3.209	3.168 E6	1.98
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	2.168	2.177	2.170	2.154	2.277	2.189 E4	2.28
23)	Toxaphene-2	8.486	8.085	8.100	8.759	9.364	8.559 E4	6.21
24)	Toxaphene-3	6.081	5.645	5.689	6.178	6.724	6.064 E4	7.21
25)	Toxaphene-4	3.574	3.398	3.384	3.627	4.009	3.599 E4	7.03
26)	Toxaphene-5	4.283	4.067	4.122	4.402	4.744	4.323 E4	6.24
27) SA	Decachlorobiphenyl	2.734	2.907	2.800	2.614	2.619	2.735 E6	4.55

Signal #2 Calibration Files

3 =PL086422.D 1 =PL086420.D 2 =PL086421.D
 4 =PL086423.D 5 =PL086424.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	0.970	0.950	0.954	0.967	1.002	0.969 E6	2.10
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	5.224	5.047	5.115	5.163	5.282	5.166 E3	1.77
23)	Toxaphene-2	6.874	6.253	6.558	6.874	7.451	6.802 E3	6.54

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
Method File : PL110423CLP.M
Title : GC Extractables
Last Update : Fri Nov 03 16:24:02 2023
Response Via : Initial Calibration

Calibration Files

3	=PL086422.D	1	=PL086420.D	2	=PL086421.D
4	=PL086423.D	5	=PL086424.D		

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	3.527	3.240	3.351	3.665	3.862	3.529	E4	7.00
25)	Toxaphene-4	2.376	2.214	2.267	2.423	2.609	2.378	E4	6.47
26)	Toxaphene-5	1.339	1.109	1.221	1.412	1.586	1.333	E4	13.67
27) SA	Decachlorobiphenyl	1.324	1.401	1.342	1.278	1.296	1.328	E6	3.59

(#) = Out of Range