

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110518\
 Data File : PL041792.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Nov 2018 19:14
 Operator : AJ\SJ
 Sample : J5764-04
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 23:48:24 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102918.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 30 01:26:00 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.472	4.104	104.3E6	55880824	16.372	18.142
28)	SA Decachlor...	8.213	9.233	95752479	39375605	14.455	15.368
Target Compounds							
2)	A alpha-BHC	3.872f	0.000	167014	0	0.017	N.D. #
6)	B beta-BHC	4.408f	0.000	6745035	0	1.790	N.D. #
9)	A Endosulfan I	5.529	0.000	2557629	0	0.344	N.D. #
11)	B alpha-Chl...	5.529f	0.000	2557629	0	0.316	N.D. #
23)	Chlordane-1	4.382	5.150	2756440	-173481	9.722	N.D. #
26)	Chlordane-4	5.529	0.000	2557629	0	3.267	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110518\
Data File : PL041792.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Nov 2018 19:14
Operator : AJ\SJ
Sample : J5764-04
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 23:48:24 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102918.M
Quant Title : GC Extractables
QLast Update : Tue Oct 30 01:26:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.472 min
Delta R.T.: -0.002 min
Response: 104296180
Conc: 16.37 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.104 min
Delta R.T.: -0.001 min
Response: 55880824
Conc: 18.14 ng/ml




#2 alpha-BHC

R.T.: 3.872 min
Delta R.T.: -0.043 min
Response: 167014
Conc: 0.02 ng/ml



#2 alpha-BHC


R.T.: 0.000 min
Exp R.T.: 4.501 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.408 min
Delta R.T.: -0.043 min
Response: 6745035
Conc: 1.79 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.957 min
Response: 0
Conc: N.D.




#9 Endosulfan I

R.T.: 5.529 min
Delta R.T.: 0.005 min
Response: 2557629
Conc: 0.34 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T. : 6.363 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 5.529 min
Delta R.T.: 0.060 min
Response: 2557629
Conc: 0.32 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T. : 6.311 min
Response: 0
Conc: N.D.




#23 Chlordane-1

R.T.: 4.382 min
Delta R.T.: 0.039 min
Response: 2756440
Conc: 9.72 ng/ml



#23 Chlordane-1


R.T.: 5.150 min
Delta R.T.: 0.036 min
Response: -173481
Conc: N.D.



#26 Chlordane-4


R.T.: 5.529 min
Delta R.T.: 0.060 min
Response: 2557629
Conc: 3.27 ng/ml

Instrument :
ECD_L
ClientSampleId :




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T. : 6.311 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.213 min
Delta R.T.: -0.006 min
Response: 95752479
Conc: 14.46 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.233 min
Delta R.T.: -0.005 min
Response: 39375605
Conc: 15.37 ng/ml