

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110519\
 Data File : PL054035.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Nov 2019 13:28
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 11/6/2019 11:41:23 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 00:17:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	4.043	591.4E6	131.0E6	54.562	52.353
28)	SA Decachlor...	8.324	9.121	494.3E6	137.8E6	50.376	48.939
Target Compounds							
2)	A alpha-BHC	3.984	4.434	979.4E6	184.7E6	56.172	53.412
3)	MA gamma-BHC...	4.270	4.720	872.2E6	170.8E6	56.006	53.029
4)	MA Heptachlor	4.577	5.229	879.8E6	162.6E6	55.743	50.466
5)	MB Aldrin	4.828	5.535	804.9E6	151.0E6	56.306m	50.640
6)	B beta-BHC	4.521	4.890	354.9E6	77571035	53.860	50.220
7)	B delta-BHC	4.729	5.107	815.4E6	144.0E6	53.571	47.656
8)	B Heptachlor...	5.274	5.921	741.1E6	140.6E6	54.899	48.204
9)	A Endosulfan I	5.616	6.278	675.4E6	135.8E6	54.947	50.105
10)	B gamma-Chl...	5.502	6.154	739.7E6	142.5E6	54.215	48.838
11)	B alpha-Chl...	5.561	6.227	714.8E6	145.4E6	53.733	49.801
12)	B 4,4'-DDE	5.724	6.383	667.6E6	141.2E6	56.181	50.120
13)	MA Dieldrin	5.860	6.534	712.1E6	142.3E6	54.635	49.877
14)	MA Endrin	6.119	6.752	638.0E6	122.6E6	55.040	50.491
15)	B Endosulfa...	6.392	6.961	593.0E6	127.2E6	53.042	49.808
16)	A 4,4'-DDD	6.241	6.873	495.8E6	106.7E6	55.710	51.195
17)	MA 4,4'-DDT	6.484	7.172	528.8E6	116.6E6	53.381	50.406
18)	B Endrin al...	6.560	7.085	476.5E6	109.7E6	51.971	49.939
19)	B Endosulfa...	6.773	7.306	520.0E6	119.8E6	51.120	48.814m
20)	A Methoxychlor	7.029	7.630	252.8E6	64382219	51.276	49.566
21)	B Endrin ke...	7.267	7.779	557.8E6	127.0E6	51.816	49.241
22)	Mirex	7.461	8.234	412.6E6	97458254	48.240	48.174

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110519\
Data File : PL054035.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Nov 2019 13:28
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
11/6/2019 11:41:23 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 00:17:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

