

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110520\
 Data File : PL062856.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Nov 2020 12:44
 Operator : DD\AJ
 Sample : L4625-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 CONTAINER-001-A-B-COMP

Manual Integrations
 APPROVED

Ankita
 11/6/2020 3:11:49 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 00:52:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 04:25:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.429f	3.950	33060789	55389667	19.220m	18.848
28)	SA Decachlor...	8.142f	8.986	35346644	45647622	22.033m	21.810

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110520\
Data File : PL062856.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Nov 2020 12:44
Operator : DD\AJ
Sample : L4625-01
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
CONTAINER-001-A-B-COMP


Manual Integrations
APPROVED

Ankita
11/6/2020 3:11:49 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 00:52:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102120.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 04:25:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.429 min
Delta R.T.: 0.017 min
Response: 33060789
Conc: 19.22 ng/ml m

Instrument :
ECD_L
ClientSampleId :
CONTAINER-001-A-B-COMP


Manual Integrations
APPROVED

Ankita
11/6/2020 3:11:49 PM




#1 Tetrachloro-m-xylene

R.T.: 3.950 min
Delta R.T.: -0.002 min
Response: 55389667
Conc: 18.85 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.142 min
Delta R.T.: 0.016 min
Response: 35346644
Conc: 22.03 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.986 min
Delta R.T.: -0.003 min
Response: 45647622
Conc: 21.81 ng/ml