

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110623\
 Data File : PL086479.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2023 17:11
 Operator : AR\AJ
 Sample : PB156920BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB156920BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 20:37:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110123.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 02 05:32:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.357	2.619	50139744	16120053	15.799	16.933
28)	SA Decachlor...	8.858	7.770	47132989	20385632	17.821	17.772

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110623\
Data File : PL086479.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2023 17:11
Operator : AR\AJ
Sample : PB156920BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB156920BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 20:37:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110123.M
Quant Title : GC Extractables
QLast Update : Thu Nov 02 05:32:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.357 min
Delta R.T.: 0.004 min
Response: 50139744
Conc: 15.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB156920BL




#1 Tetrachloro-m-xylene

R.T.: 2.619 min
Delta R.T.: 0.000 min
Response: 16120053
Conc: 16.93 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.858 min
Delta R.T.: 0.003 min
Response: 47132989
Conc: 17.82 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.770 min
Delta R.T.: 0.000 min
Response: 20385632
Conc: 17.77 ng/ml