

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110719\
 Data File : PL054068.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2019 14:39
 Operator : SG\AJ
 Sample : K5723-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 7-C

Manual Integrations
 APPROVED

Ankita
 11/8/2019 8:46:40 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 00:18:28 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	4.044	203.4E6	47002800	18.762	18.783
28)	SA Decachlor...	8.324	9.122	166.5E6	48447983	16.969	17.204m
Target Compounds							
16)	A 4,4'-DDD	6.244	6.873	9282455	8679116	1.043	4.165m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110719\
Data File : PL054068.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2019 14:39
Operator : SG\AJ
Sample : K5723-03
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
7-C


Manual Integrations
APPROVED

Ankita
11/8/2019 8:46:40 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 00:18:28 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 203365158
Conc: 18.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
7-C


Manual Integrations
APPROVED

Ankita
11/8/2019 8:46:40 AM




#1 Tetrachloro-m-xylene

R.T.: 4.044 min
Delta R.T.: -0.002 min
Response: 47002800
Conc: 18.78 ng/ml




#16 4,4'-DDD

R.T.: 6.244 min
Delta R.T.: -0.005 min
Response: 9282455
Conc: 1.04 ng/ml



#16 4,4'-DDD

R.T.: 6.873 min
Delta R.T.: -0.004 min
Response: 8679116
Conc: 4.16 ng/ml m




#28 Decachlorobiphenyl

R.T.: 8.324 min
Delta R.T.: -0.010 min
Response: 166494151
Conc: 16.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
7-C

Manual Integrations
APPROVED

Ankita
11/8/2019 8:46:40 AM



#28 Decachlorobiphenyl

R.T.: 9.122 min
Delta R.T.: -0.005 min
Response: 48447983
Conc: 17.20 ng/ml m