

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL110722\  
 Data File : PL078857.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 07 Nov 2022 15:03  
 Operator : AR\AJ  
 Sample : N5436-05  
 Misc :  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 ECD\_L  
 ClientSampleId :  
 250FERRYBLVD-TFS-05

Manual Integrations  
 APPROVED

Reviewed By :Abdul Mirza 11/08/2022  
 Supervised By :Ankita Jodhani 11/10/2022

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Nov 09 17:23:04 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL102122.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Oct 21 07:08:41 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2   | ng/ml   | ng/ml    |
|-----------------------------|-----------------|-------|-------|----------|----------|---------|----------|
| -----                       |                 |       |       |          |          |         |          |
| System Monitoring Compounds |                 |       |       |          |          |         |          |
| 1)                          | SA Tetrachlo... | 3.323 | 2.570 | 45745581 | 25872326 | 23.988  | 10.133 # |
| 28)                         | SA Decachlor... | 8.774 | 7.639 | 295.5E6  | 365.7E6  | 196.416 | 193.093  |
| Target Compounds            |                 |       |       |          |          |         |          |
| 13)                         | MA Dieldrin     | 6.118 | 5.105 | 168.5E6  | 134.0E6  | 78.131  | 45.224m# |
| -----                       |                 |       |       |          |          |         |          |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL110722\  
Data File : PL078857.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 07 Nov 2022 15:03  
Operator : AR\AJ  
Sample : N5436-05  
Misc :  
ALS Vial : 13 Sample Multiplier: 1



Instrument :  
ECD\_L  
ClientSampleId :  
250FERRYBLVD-TFS-05


Manual Integrations  
APPROVED

Reviewed By :Abdul Mirza 11/08/2022  
Supervised By :Ankita Jodhani 11/10/2022

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 09 17:23:04 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL102122.M  
Quant Title : GC Extractables  
QLast Update : Fri Oct 21 07:08:41 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






# #1 Tetrachloro-m-xylene

R.T.: 3.323 min  
Delta R.T.: 0.000 min  
Response: 45745581  
Conc: 23.99 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :  
250FERRYBLVD-TFS-05


Manual Integrations  
APPROVED

Reviewed By :Abdul Mirza 11/08/2022  
Supervised By :Ankita Jodhani 11/10/2022




# #1 Tetrachloro-m-xylene

R.T.: 2.570 min  
Delta R.T.: -0.004 min  
Response: 25872326  
Conc: 10.13 ng/ml




# #13 Dieldrin

R.T.: 6.118 min  
Delta R.T.: -0.003 min  
Response: 168490530  
Conc: 78.13 ng/ml



# #13 Dieldrin

R.T.: 5.105 min  
Delta R.T.: 0.003 min  
Response: 134002461  
Conc: 45.22 ng/ml m




#28 Decachlorobiphenyl

R.T.: 8.774 min  
Delta R.T.: -0.003 min  
Response: 295479628  
Conc: 196.42 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :  
250FERRYBLVD-TFS-05

Manual Integrations  
APPROVED

Reviewed By :Abdul Mirza 11/08/2022  
Supervised By :Ankita Jodhani 11/10/2022



#28 Decachlorobiphenyl

R.T.: 7.639 min  
Delta R.T.: -0.005 min  
Response: 365737435  
Conc: 193.09 ng/ml