

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110819\
 Data File : PL054128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2019 09:02
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 11/11/2019 1:18:51 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 22:03:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	4.043	563.1E6	127.3E6	51.954	50.879
28)	SA Decachlor...	8.324	9.123	471.4E6	134.2E6	48.047	47.672
Target Compounds							
2)	A alpha-BHC	3.984	4.434	942.8E6	180.0E6	54.072	52.054
3)	MA gamma-BHC...	4.270	4.720	835.8E6	166.6E6	53.668	51.737
4)	MA Heptachlor	4.577	5.229	830.3E6	156.4E6	52.609	48.544
5)	MB Aldrin	4.827	5.535	776.2E6	147.6E6	54.299m	49.491
6)	B beta-BHC	4.521	4.890	344.4E6	77633351	52.256	50.260
7)	B delta-BHC	4.729	5.107	782.8E6	151.0E6	51.431	49.995
8)	B Heptachlor...	5.274	5.921	728.6E6	145.4E6	53.973	49.864
9)	A Endosulfan I	5.616	6.278	663.7E6	132.8E6	53.998	48.992
10)	B gamma-Chl...	5.503	6.154	726.5E6	142.5E6	53.248	48.848m
11)	B alpha-Chl...	5.561	6.228	693.6E6	141.8E6	52.135	48.581
12)	B 4,4'-DDE	5.724	6.384	631.8E6	136.6E6	53.170	48.507
13)	MA Dieldrin	5.860	6.535	676.5E6	138.9E6	51.905	48.674
14)	MA Endrin	6.119	6.753	583.0E6	113.7E6	50.297	46.848
15)	B Endosulfa...	6.392	6.962	565.0E6	125.8E6	50.537	49.251
16)	A 4,4'-DDD	6.242	6.874	485.7E6	108.0E6	54.576	51.846
17)	MA 4,4'-DDT	6.484	7.173	481.5E6	107.1E6	48.610	46.303
18)	B Endrin al...	6.560	7.086	460.0E6	107.3E6	50.167	48.828
19)	B Endosulfa...	6.774	7.307	501.9E6	117.0E6	49.343	47.686m
20)	A Methoxychlor	7.029	7.631	235.5E6	59510626	47.760	45.816
21)	B Endrin ke...	7.267	7.780	545.7E6	127.1E6	50.695	49.298
22)	Mirex	7.460	8.235	393.9E6	96311086	46.053	47.607

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110819\
Data File : PL054128.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2019 09:02
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
11/11/2019 1:18:51 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 22:03:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

