

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110821\
 Data File : PL070836.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2021 17:12
 Operator : AR\AJ
 Sample : M4475-14
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 00:47:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110321.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 04 19:28:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.568	5.501	9373232	25241496	19.137	20.019
28) SA Decachlor...	10.330	11.391	9991116	21373969	14.622	15.982
Target Compounds						
7) B delta-BHC	0.000	7.022f	0	742049	N.D.	0.444 #
12) B 4,4'-DDE	0.000	8.491	0	539985	N.D.	0.366 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110821\
Data File : PL070836.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2021 17:12
Operator : AR\AJ
Sample : M4475-14
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 00:47:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110321.M
Quant Title : GC Extractables
QLast Update : Thu Nov 04 19:28:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.568 min
Delta R.T.: 0.000 min
Response: 9373232
Conc: 19.14 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 5.501 min
Delta R.T.: 0.000 min
Response: 25241496
Conc: 20.02 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 6.338 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 7.022 min
Delta R.T.: 0.024 min
Response: 742049
Conc: 0.44 ng/ml



#12 4,4'-DDE


R.T.: 0.000 min
Exp R.T.: 7.567 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#12 4,4'-DDE

R.T.: 8.491 min
Delta R.T.: -0.010 min
Response: 539985
Conc: 0.37 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.330 min
Delta R.T.: 0.000 min
Response: 9991116
Conc: 14.62 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.391 min
Delta R.T.: -0.001 min
Response: 21373969
Conc: 15.98 ng/ml