

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110823\
 Data File : PL086533.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2023 12:36
 Operator : AR\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL110823CHLOR

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/08/2023
 Supervised By : Ankita Jodhani 11/09/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 13:01:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110823.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 08 13:01:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.351	2.617	171.3E6	64539166	54.824	55.070
28)	SA Decachlor...	8.852	7.768	152.9E6	63014502	53.839	53.341
Target Compounds							
23)	Chlordane-1	4.508	3.605	76782484	18683611	548.797	533.274
24)	Chlordane-2	5.041	4.183	81250048	24202946	536.053	535.856
25)	Chlordane-3	5.752	4.812	319.8E6	65951092	552.868	559.324
26)	Chlordane-4	5.834	4.873	391.7E6	70438706	547.829	551.704
27)	Chlordane-5	6.681	5.544	65575725	19954028	503.969m	549.604

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110823\
Data File : PL086533.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2023 12:36
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL110823CHLOR

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/08/2023
Supervised By : Ankita Jodhani 11/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 13:01:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110823.M
Quant Title : GC Extractables
QLast Update : Wed Nov 08 13:01:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

