

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110824\
 Data File : PL092939.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2024 17:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 22:26:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 28 18:58:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.779	122.3E6	143.1E6	49.896	52.590
2)	SA Decachlor...	9.060	7.919	81818276	118.0E6	42.513	43.241
Target Compounds							
2)	A alpha-BHC	3.998	3.281	169.4E6	213.8E6	49.956	53.579
3)	MA gamma-BHC...	4.330	3.612	160.6E6	204.5E6	49.256	52.999
4)	MA Heptachlor	4.919	3.951	137.2E6	189.2E6	45.673	50.122
5)	MB Aldrin	5.261	4.230	140.4E6	193.6E6	46.780	52.892
6)	B beta-BHC	4.528	3.911	70484349	86215676	48.763	52.371
7)	B delta-BHC	4.776	4.140	153.2E6	206.0E6	48.952	53.608
8)	B Heptachlo...	5.687	4.733	126.3E6	174.1E6	45.726	52.106
9)	A Endosulfan I	6.073	5.103	112.9E6	155.2E6	45.124	50.825
10)	B gamma-Chl...	5.943	4.983	123.0E6	173.1E6	46.240	51.494
11)	B alpha-Chl...	6.023	5.047	121.1E6	169.0E6	45.720	50.823
12)	B 4,4'-DDE	6.195	5.236	109.1E6	167.1E6	46.044	51.880
13)	MA Dieldrin	6.348	5.367	119.2E6	170.2E6	45.230	50.961
14)	MA Endrin	6.578	5.643	93686170	140.2E6	41.151	48.462
15)	B Endosulfa...	6.798	5.938	103.2E6	140.8E6	43.277	49.705
16)	A 4,4'-DDD	6.713	5.791	93092639	133.7E6	48.498	53.696
17)	MA 4,4'-DDT	7.027	6.041	84240574	119.2E6	40.701	44.440
18)	B Endrin al...	6.928	6.117	81976293	111.6E6	43.651	48.377
19)	B Endosulfa...	7.163	6.340	94109390	129.3E6	43.362	47.934
20)	A Methoxychlor	7.503	6.617	46315609	62540606	40.628	43.795
21)	B Endrin ke...	7.647	6.845	105.5E6	148.7E6	43.553	48.443
22)	Mirex	8.121	7.027	80969128	116.1E6	40.858	44.514

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110824\
Data File : PL092939.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2024 17:49
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 22:26:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
Quant Title : GC Extractables
QLast Update : Mon Oct 28 18:58:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

