

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110918\
 Data File : PL042037.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2018 05:33
 Operator : AJ\SJ
 Sample : J5884-15
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 02:49:20 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110918.M
 Quant Title : GC Extractables
 QLast Update : Fri Nov 09 12:38:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.469	4.103	45707686	23485885	7.402	7.927
28)	SA Decachlor...	8.211	9.236	27482265	9232365	4.012	3.620
Target Compounds							
6)	B beta-BHC	4.403f	0.000	4939362	0	1.338	N.D. #
17)	MA 4,4'-DDT	0.000	7.261	0	7636850	N.D.	2.832 #
18)	B Endrin al...	0.000	7.171	0	18904395	N.D.	7.627 #
23)	Chlordane-1	4.403	0.000	4939362	0	16.551	N.D. #
27)	Chlordane-5	0.000	7.171	0	18904395	N.D.	161.400 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110918\
Data File : PL042037.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2018 05:33
Operator : AJ\SJ
Sample : J5884-15
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 02:49:20 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110918.M
Quant Title : GC Extractables
QLast Update : Fri Nov 09 12:38:40 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.469 min
Delta R.T.: 0.000 min
Response: 45707686
Conc: 7.40 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.103 min
Delta R.T.: 0.001 min
Response: 23485885
Conc: 7.93 ng/ml




#6 beta-BHC

R.T.: 4.403 min
Delta R.T.: -0.042 min
Response: 4939362
Conc: 1.34 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 4.954 min
Response: 0
Conc: N.D.



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T.: 6.381 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.261 min
Delta R.T.: 0.008 min
Response: 7636850
Conc: 2.83 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T.: 6.457 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 7.171 min
Delta R.T.: 0.003 min
Response: 18904395
Conc: 7.63 ng/ml



#23 Chlordane-1


R.T.: 4.403 min
Delta R.T.: 0.066 min
Response: 4939362
Conc: 16.55 ng/ml

Instrument :
ECD_L
ClientSampleId :




#23 Chlordane-1

R.T.: 0.000 min
Exp R.T. : 5.112 min
Response: 0
Conc: N.D.




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.296 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 7.171 min
Delta R.T.: 0.052 min
Response: 18904395
Conc: 161.40 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.211 min
Delta R.T.: 0.003 min
Response: 27482265
Conc: 4.01 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.236 min
Delta R.T.: 0.005 min
Response: 9232365
Conc: 3.62 ng/ml