

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110922\
 Data File : PL078980.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2022 21:05
 Operator : AR\AJ
 Sample : N5509-10
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-11

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 09:09:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.320	2.568	27131850	35327932	14.227	13.837
28)	SA Decachlor...	8.767	7.634	16082927	21646613	10.691	11.428
Target Compounds							
10)	B gamma-Chl...	5.710	4.718	2039246	1983476	0.903	0.647 #
11)	B alpha-Chl...	5.788	4.778	3556119	3022237	1.537	1.014 #
12)	B 4,4'-DDE	5.969	4.974	944704	1135632	0.472	0.417



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110922\
Data File : PL078980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2022 21:05
Operator : AR\AJ
Sample : N5509-10
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-11

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 09:09:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.320 min
Delta R.T.: -0.003 min
Response: 27131850
Conc: 14.23 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-11




#1 Tetrachloro-m-xylene

R.T.: 2.568 min
Delta R.T.: -0.005 min
Response: 35327932
Conc: 13.84 ng/ml




#10 gamma-Chlordane

R.T.: 5.710 min
Delta R.T.: -0.004 min
Response: 2039246
Conc: 0.90 ng/ml



#10 gamma-Chlordane


R.T.: 4.718 min
Delta R.T.: -0.007 min
Response: 1983476
Conc: 0.65 ng/ml



#11 alpha-Chlordane


R.T.: 5.788 min
Delta R.T.: -0.003 min
Response: 3556119
Conc: 1.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-11




#11 alpha-Chlordane

R.T.: 4.778 min
Delta R.T.: -0.009 min
Response: 3022237
Conc: 1.01 ng/ml




#12 4,4'-DDE

R.T.: 5.969 min
Delta R.T.: -0.005 min
Response: 944704
Conc: 0.47 ng/ml



#12 4,4'-DDE


R.T.: 4.974 min
Delta R.T.: -0.008 min
Response: 1135632
Conc: 0.42 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.767 min
Delta R.T.: -0.009 min
Response: 16082927
Conc: 10.69 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-11



#28 Decachlorobiphenyl

R.T.: 7.634 min
Delta R.T.: -0.010 min
Response: 21646613
Conc: 11.43 ng/ml