

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : PL110923CLP.M
 Title : GC Extractables
 Last Update : Wed Nov 08 21:00:47 2023
 Response Via : Initial Calibration

Calibration Files

3 =PL086540.D 1 =PL086538.D 2 =PL086539.D
 4 =PL086541.D 5 =PL086542.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.483	3.429	3.444	3.609	3.619	3.517	E6	2.59
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	2.288	2.237	2.361	2.464	2.518	2.374	E4	4.96
23)	Toxaphene-2	0.940	0.857	0.910	0.997	1.026	0.946	E5	7.17
24)	Toxaphene-3	6.773	6.198	6.514	7.354	7.562	6.880	E4	8.28
25)	Toxaphene-4	4.067	3.598	3.847	4.225	4.422	4.032	E4	7.97
26)	Toxaphene-5	4.937	4.567	4.707	5.236	5.533	4.996	E4	7.86
27) SA	Decachlorobiphenyl	3.057	3.079	3.091	3.075	2.999	3.060	E6	1.18

Signal #2 Calibration Files

3 =PL086540.D 1 =PL086538.D 2 =PL086539.D
 4 =PL086541.D 5 =PL086542.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.058	0.991	1.029	1.081	1.104	1.053	E6	4.20
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	5.791	5.522	5.733	4.824	4.847	5.343	E3	8.88

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : PL110923CLP.M
 Title : GC Extractables
 Last Update : Wed Nov 08 21:00:47 2023
 Response Via : Initial Calibration

Calibration Files

3 =PL086540.D 1 =PL086538.D 2 =PL086539.D
 4 =PL086541.D 5 =PL086542.D

	Compound	3	1	2	4	5	Avg		%RSD
23)	Toxaphene-2	6.885	6.221	6.857	6.673	7.635	6.854	E3	7.45
24)	Toxaphene-3	3.562	3.195	3.357	3.795	3.974	3.577	E4	8.84
25)	Toxaphene-4	2.375	2.135	2.311	2.525	2.676	2.404	E4	8.58
26)	Toxaphene-5	1.353	1.113	1.292	1.506	1.635	1.380	E4	14.52
27) SA	Decachlorobiphenyl	1.393	1.412	1.424	1.404	1.400	1.407	E6	0.87

(#) = Out of Range