

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111119\
 Data File : PL054181.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2019 18:06
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 11/12/2019 11:50:17 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 00:28:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	4.044	567.5E6	130.9E6	52.359	52.322
28)	SA Decachlor...	8.324	9.124	444.5E6	130.0E6	45.304	46.179
Target Compounds							
2)	A alpha-BHC	3.984	4.435	939.6E6	185.7E6	53.889	53.707
3)	MA gamma-BHC...	4.270	4.721	833.1E6	170.0E6	53.492	52.793
4)	MA Heptachlor	4.576	5.231	835.3E6	158.6E6	52.923	49.215
5)	MB Aldrin	4.828	5.536	772.4E6	145.7E6	54.032	48.851
6)	B beta-BHC	4.521	4.891	339.2E6	77756129	51.475	50.340
7)	B delta-BHC	4.729	5.109	797.4E6	152.5E6	52.386	50.488
8)	B Heptachlo...	5.274	5.922	696.2E6	133.8E6	51.574	45.876
9)	A Endosulfan I	5.615	6.279	623.1E6	125.6E6	50.688	46.322
10)	B gamma-Chl...	5.502	6.156	697.2E6	135.4E6	51.100	46.420
11)	B alpha-Chl...	5.560	6.229	672.0E6	135.5E6	50.514	46.422
12)	B 4,4'-DDE	5.724	6.385	630.9E6	131.6E6	53.094	46.728
13)	MA Dieldrin	5.860	6.536	654.8E6	133.5E6	50.236	46.771
14)	MA Endrin	6.119	6.754	579.4E6	114.9E6	49.988	47.328
15)	B Endosulfa...	6.392	6.963	525.7E6	115.5E6	47.020	45.222
16)	A 4,4'-DDD	6.242	6.875	456.6E6	100.6E6	51.309	48.250
17)	MA 4,4'-DDT	6.484	7.173	469.4E6	106.8E6	47.383	46.177
18)	B Endrin al...	6.561	7.087	416.2E6	99405424	45.394	45.251
19)	B Endosulfa...	6.774	7.310	478.9E6	110.5E6	47.086	45.039
20)	A Methoxychlor	7.030	7.632	231.5E6	58440507	46.953	44.992
21)	B Endrin ke...	7.267	7.782	504.2E6	117.8E6	46.839m	45.684
22)	Mirex	7.459	8.236	377.3E6	92005395	44.114m	45.478

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111119\
Data File : PL054181.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2019 18:06
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
11/12/2019 11:50:17 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 00:28:12 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

