

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111122\
 Data File : PL079069.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2022 04:10
 Operator : AR\AJ
 Sample : N5561-01
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SB-25

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 04:37:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 07:08:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.321	2.569	22943670	29703389	12.031	11.634
28)	SA Decachlor...	8.773	7.636	15984228	22687470	10.625	11.978
Target Compounds							
11)	B alpha-Chl...	5.791	4.778	7713980	3714123	3.334	1.246 #
12)	B 4,4'-DDE	5.973	4.975	12039234	17508886	6.016	6.431
16)	A 4,4'-DDD	6.496	5.525	4181379	4726426	2.508m	2.152
17)	MA 4,4'-DDT	6.806	5.771	2456405	5858473	1.371m	2.701 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111122\
Data File : PL079069.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2022 04:10
Operator : AR\AJ
Sample : N5561-01
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SB-25

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 04:37:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102122.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 07:08:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.321 min
Delta R.T.: -0.002 min
Response: 22943670
Conc: 12.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-25




#1 Tetrachloro-m-xylene

R.T.: 2.569 min
Delta R.T.: -0.005 min
Response: 29703389
Conc: 11.63 ng/ml




#11 alpha-Chlordane

R.T.: 5.791 min
Delta R.T.: 0.000 min
Response: 7713980
Conc: 3.33 ng/ml



#11 alpha-Chlordane


R.T.: 4.778 min
Delta R.T.: -0.009 min
Response: 3714123
Conc: 1.25 ng/ml



#12 4,4'-DDE


R.T.: 5.973 min
Delta R.T.: -0.001 min
Response: 12039234
Conc: 6.02 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-25




#12 4,4'-DDE

R.T.: 4.975 min
Delta R.T.: -0.007 min
Response: 17508886
Conc: 6.43 ng/ml




#16 4,4'-DDD

R.T.: 6.496 min
Delta R.T.: -0.003 min
Response: 4181379
Conc: 2.51 ng/ml m



#16 4,4'-DDD


R.T.: 5.525 min
Delta R.T.: -0.007 min
Response: 4726426
Conc: 2.15 ng/ml



#17 4,4'-DDT


R.T.: 6.806 min
Delta R.T.: -0.003 min
Response: 2456405
Conc: 1.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-25




#17 4,4'-DDT

R.T.: 5.771 min
Delta R.T.: -0.008 min
Response: 5858473
Conc: 2.70 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.004 min
Response: 15984228
Conc: 10.63 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.636 min
Delta R.T.: -0.008 min
Response: 22687470
Conc: 11.98 ng/ml