

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111218\
 Data File : PL042047.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2018 13:52
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 11/13/2018 11:48:53 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 03:33:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110918.M
 Quant Title : GC Extractables
 QLast Update : Fri Nov 09 12:38:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.469	4.102	296.5E6	143.3E6	48.019	48.361
28)	SA Decachlor...	8.207	9.230	281.4E6	111.5E6	41.072	43.712
Target Compounds							
2)	A alpha-BHC	3.909	4.497	463.3E6	208.8E6	48.858	47.023
3)	MA gamma-BHC...	4.192	4.786	421.4E6	192.2E6	48.829	46.619
4)	MA Heptachlor	4.488	5.301	432.3E6	190.5E6	48.551	47.316
5)	MB Aldrin	4.734	5.610	397.8E6	178.3E6	48.072	47.415
6)	B beta-BHC	4.445	4.954	174.7E6	80228443	47.339	48.817
7)	B delta-BHC	4.647	5.175	398.9E6	183.0E6	49.261	46.288
8)	B Heptachlor...	5.177	5.997	371.8E6	164.6E6	47.132	45.810
9)	A Endosulfan I	5.515	6.357	341.4E6	148.1E6	46.848	45.417
10)	B gamma-Chl...	5.401	6.232	378.6E6	163.4E6	46.981m	46.111
11)	B alpha-Chl...	5.460	6.305	371.1E6	159.0E6	46.564	45.710
12)	B 4,4'-DDE	5.627	6.460	329.5E6	157.3E6	48.158	46.067
13)	MA Dieldrin	5.758	6.616	366.6E6	156.1E6	47.118	46.618
14)	MA Endrin	6.014	6.835	322.0E6	128.4E6	45.270	43.631
15)	B Endosulfa...	6.288	7.043	312.9E6	130.0E6	45.843	43.727
16)	A 4,4'-DDD	6.142	6.951	267.4E6	121.1E6	48.375	45.934
17)	MA 4,4'-DDT	6.380	7.251	277.6E6	119.3E6	48.801	44.261
18)	B Endrin al...	6.456	7.165	253.9E6	108.4E6	45.401	43.734m
19)	B Endosulfa...	6.669	7.390	288.4E6	119.0E6	45.224	43.741
20)	A Methoxychlor	6.927	7.707	149.2E6	58610591	46.841	42.654
21)	B Endrin ke...	7.158	7.865	331.2E6	125.0E6	45.963	43.173

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111218\
Data File : PL042047.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2018 13:52
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
11/13/2018 11:48:53 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 03:33:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110918.M
Quant Title : GC Extractables
QLast Update : Fri Nov 09 12:38:40 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

