

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111219\
 Data File : PL054205.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2019 16:36
 Operator : SG\AJ
 Sample : K5789-03MS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-17-SMS

Manual Integrations
 APPROVED

Ankita
 11/13/2019 10:45:27 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 00:32:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 13:56:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	4.043	237.6E6	59708237	21.917	23.861
28)	SA Decachlor...	8.325	9.124	185.1E6	54162588	18.867	19.234
Target Compounds							
2)	A alpha-BHC	3.984	4.436	1010.0E6	212.7E6	57.922	61.537
3)	MA gamma-BHC...	4.270	4.721	909.3E6	198.9E6	58.390	61.751
4)	MA Heptachlor	4.577	5.230	904.9E6	181.9E6	57.337	56.441
5)	MB Aldrin	4.828	5.535	826.6E6	169.6E6	57.824	56.868
6)	B beta-BHC	4.521	4.891	366.8E6	89212910	55.666	57.757
7)	B delta-BHC	4.730	5.109	764.3E6	145.6E6	50.213	48.190
8)	B Heptachlor...	5.274	5.922	758.1E6	143.3E6	56.157	49.124
9)	A Endosulfan I	5.616	6.279	661.3E6	142.2E6	53.796	52.460
10)	B gamma-Chl...	5.502	6.155	736.5E6	154.0E6	53.978	52.768m
11)	B alpha-Chl...	5.561	6.229	704.9E6	150.9E6	52.983	51.708
12)	B 4,4'-DDE	5.724	6.385	669.6E6	148.6E6	56.355	52.777
13)	MA Dieldrin	5.860	6.536	703.3E6	150.6E6	53.959	52.788
14)	MA Endrin	6.119	6.754	639.9E6	130.5E6	55.207	53.768
15)	B Endosulfa...	6.392	6.964	570.4E6	132.0E6	51.020	51.679
16)	A 4,4'-DDD	6.242	6.875	490.3E6	114.6E6	55.086	54.997
17)	MA 4,4'-DDT	6.484	7.174	524.7E6	123.1E6	52.966	53.252
18)	B Endrin al...	6.561	7.087	451.7E6	111.9E6	49.259	50.925
19)	B Endosulfa...	6.774	7.310	501.6E6	120.3E6	49.317	49.048
20)	A Methoxychlor	7.029	7.632	260.0E6	67356832	52.743	51.856
21)	B Endrin ke...	7.268	7.782	542.5E6	133.7E6	50.392	51.872
22)	Mirex	7.461	8.236	441.7E6	106.8E6	51.641	52.777

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL111219\
Data File : PL054205.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2019 16:36
Operator : SG\AJ
Sample : K5789-03MS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
TP-17-SMS

Manual Integrations
APPROVED

Ankita
11/13/2019 10:45:27 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 00:32:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL101919.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 13:56:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

